

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF061818\
 Data File : VF059182.D
 Acq On : 18 Jun 2018 18:16
 Operator : VA/AP
 Sample : J3567-07
 Misc : 5.31g/5mL/MSVOA-F/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 TP-5-D

Manual Integrations
 APPROVED

apatel
 6/19/2018 3:54:08 PM

Quant Time: Jun 19 04:10:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F060618S.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 07 10:52:00 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.03	168	258745	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.75	114	405287	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.91	117	391004	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.63	152	226568	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.02	65	201887	64.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	128.68%	
35) Dibromofluoromethane	4.29	113	174852	61.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.98%	
50) Toluene-d8	7.71	98	424976	50.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.74%	
62) 4-Bromofluorobenzene	11.50	95	210543	49.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.26%	
Target Compounds						
16) Acetone	2.33	43	18325	24.458	ug/l	97
20) Methylene Chloride	2.31	84	7908m	4.408	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF061818\
Data File : VF059182.D
Acq On : 18 Jun 2018 18:16
Operator : VA/AP
Sample : J3567-07
Misc : 5.31g/5mL/MSVOA-F/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
TP-5-D

Manual Integrations
APPROVED

apatel
6/19/2018 3:54:08 PM

Quant Time: Jun 19 04:10:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F060618S.M
Quant Title : SW846 8260
QLast Update : Thu Jun 07 10:52:00 2018
Response via : Initial Calibration

