

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF062218\
 Data File : VF059295.D
 Acq On : 23 Jun 2018 1:10
 Operator : VA/AP
 Sample : J3605-05
 Misc : 5.34g/5mL/MSVOA-F/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 S-4(3-3.5)

Manual Integrations
 APPROVED

apatel
 6/24/2018 11:35:02 AM

Quant Time: Jun 23 05:14:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F062118S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 22 05:22:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.04	168	181958	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.76	114	285076	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.91	117	259497	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.63	152	132790	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.02	65	143108	54.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.58%	
35) Dibromofluoromethane	4.29	113	38231	15.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	31.88%	
50) Toluene-d8	7.70	98	270924	44.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.82%	
62) 4-Bromofluorobenzene	11.50	95	127763	37.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.54%	
Target Compounds						
16) Acetone	2.34	43	17560	28.45	ug/l	99
20) Methylene Chloride	2.28	84	13906m	9.58	ug/l	
27) cis-1,2-Dichloroethene	3.63	96	29795	14.57	ug/l	98
44) Trichloroethene	5.67	130	29049	13.19	ug/l	92
64) Tetrachloroethene	8.30	164	1207593	605.93	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF062218\
Data File : VF059295.D
Acq On : 23 Jun 2018 1:10
Operator : VA/AP
Sample : J3605-05
Misc : 5.34g/5mL/MSVOA-F/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
S-4(3-3.5)

Manual Integrations
APPROVED

apatel
6/24/2018 11:35:02 AM

Quant Time: Jun 23 05:14:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F062118S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 22 05:22:31 2018
Response via : Initial Calibration

