

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF062218\
 Data File : VF059296.D
 Acq On : 23 Jun 2018 1:38
 Operator : VA/AP
 Sample : J3605-06
 Misc : 6.59g/5mL/MSVOA-F/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 S-5(6.5-7)

Manual Integrations
 APPROVED

apatel
 6/24/2018 11:35:18 AM

Quant Time: Jun 23 05:16:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F062118S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 22 05:22:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.03	168	172475	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	5.75	114	272686	50.00	ug/l	-0.01
63) Chlorobenzene-d5	9.91	117	283990	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.62	152	178097	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.01	65	132471	53.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.02%	
35) Dibromofluoromethane	4.29	113	118103	51.48	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	102.96%	
50) Toluene-d8	7.71	98	278352	47.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.40%	
62) 4-Bromofluorobenzene	11.50	95	157615	48.71	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	97.42%	
Target Compounds						
20) Methylene Chloride	2.26	84	14364m	10.62	ug/l	Qvalue
64) Tetrachloroethene	8.30	164	20413	9.36	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF062218\
Data File : VF059296.D
Acq On : 23 Jun 2018 1:38
Operator : VA/AP
Sample : J3605-06
Misc : 6.59g/5mL/MSVOA-F/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
S-5(6.5-7)

Manual Integrations
APPROVED

apatel
6/24/2018 11:35:18 AM

Quant Time: Jun 23 05:16:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F062118S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 22 05:22:31 2018
Response via : Initial Calibration

