

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF062718\
 Data File : VF059336.D
 Acq On : 27 Jun 2018 13:56
 Operator : VA/AP
 Sample : J3684-07
 Misc : 6.12g/5mL/MSVOA-F/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 TP-2-S

Quant Time: Jun 27 17:04:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F062518S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 26 08:45:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.03	168	237133	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.75	114	372836	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.90	117	361806	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.63	152	216609	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.02	65	148101	48.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.68%	
35) Dibromofluoromethane	4.29	113	134907	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
50) Toluene-d8	7.70	98	315983	46.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.70%	
62) 4-Bromofluorobenzene	11.50	95	175364	46.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.18%	
Target Compounds						
16) Acetone	2.33	43	49528	56.85	ug/l	96
20) Methylene Chloride	2.27	84	14448	1.20	ug/l	85
40) Benzene	4.80	78	20202	2.80	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF062718\
Data File : VF059336.D
Acq On : 27 Jun 2018 13:56
Operator : VA/AP
Sample : J3684-07
Misc : 6.12g/5mL/MSVOA-F/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
TP-2-S

Quant Time: Jun 27 17:04:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F062518S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 26 08:45:17 2018
Response via : Initial Calibration

