


Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF062918\
 Data File : VF059375.D
 Acq On : 29 Jun 2018 12:53
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00µ/5mL/MSVOA-F/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F062518S.M
 Title : SW846 8260
 Last Update : Tue Jun 26 08:45:17 2018



AutoFind: Scans 1112, 1113, 1114; Background Corrected with Scan 1107

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	27285	PASS
75	95	30	60	47.7	66354	PASS
95	95	100	100	100.0	139008	PASS
96	95	5	9	6.8	9449	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	82.0	113968	PASS
175	174	5	9	7.7	8732	PASS
176	174	95	101	97.5	111152	PASS
177	176	5	9	6.5	7216	PASS