

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF070218\
 Data File : VF059391.D
 Acq On : 2 Jul 2018 17:09
 Operator : VA/AP
 Sample : J3802-05
 Misc : 5.83g/5mL/MSVOA-F/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 CF-5(6.0)

Quant Time: Jul 03 11:50:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F062518S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 26 08:45:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.02	168	99425	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.74	114	152043	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.90	117	98081	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.62	152	73529	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.99	65	85999	66.95	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	133.90%	
35) Dibromofluoromethane	4.27	113	75106	67.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	135.22%	
50) Toluene-d8	7.69	98	98169	35.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.38%	
62) 4-Bromofluorobenzene	11.50	95	80112	51.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.26%	
Target Compounds						
16) Acetone	2.33	43	22814	63.50	ug/l	96
20) Methylene Chloride	2.27	84	18417	21.33	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF070218\
Data File : VF059391.D
Acq On : 2 Jul 2018 17:09
Operator : VA/AP
Sample : J3802-05
Misc : 5.83g/5mL/MSVOA-F/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
CF-5(6.0)

Quant Time: Jul 03 11:50:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F062518S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 26 08:45:17 2018
Response via : Initial Calibration

