

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF070218\
 Data File : VF059393.D
 Acq On : 2 Jul 2018 18:04
 Operator : VA/AP
 Sample : J3802-07
 Misc : 6.50u/5mL/MSVOA-F/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 CF-7(3.0)

Quant Time: Jul 03 02:05:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F062518S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 26 08:45:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.03	168	175099	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.75	114	297920	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.90	117	293645	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.62	152	173146	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.00	65	169776	75.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	150.10%	
35) Dibromofluoromethane	4.28	113	145013	66.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	133.24%	
50) Toluene-d8	7.69	98	345021	64.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	128.04%	
62) 4-Bromofluorobenzene	11.49	95	180257	59.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.58%	
Target Compounds						
16) Acetone	2.33	43	68295	115.39	ug/l	95
25) 2-Butanone	4.51	43	6547	6.90	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF070218\
Data File : VF059393.D
Acq On : 2 Jul 2018 18:04
Operator : VA/AP
Sample : J3802-07
Misc : 6.50g/5mL/MSVOA-F/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
CF-7(3.0)

Quant Time: Jul 03 02:05:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F062518S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 26 08:45:17 2018
Response via : Initial Calibration

