

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF070218\
 Data File : VF059399.D
 Acq On : 2 Jul 2018 20:51
 Operator : VA/AP
 Sample : J3802-05RE
 Misc : 6.01g/5mL/MSVOA-F/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 CF-5(6.0)RE

Quant Time: Jul 03 11:49:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F062518S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 26 08:45:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.01	168	151349	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	5.74	114	227500	50.00	ug/l	-0.01
63) Chlorobenzene-d5	9.89	117	200939	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	12.61	152	118464	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.00	65	109692	56.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.20%	
35) Dibromofluoromethane	4.28	113	103041	61.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	123.98%	
50) Toluene-d8	7.69	98	245402	59.63	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	119.26%	
62) 4-Bromofluorobenzene	11.49	95	123782	53.32	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	106.64%	
Target Compounds						
16) Acetone	2.33	43	56080	109.09	ug/l	Qvalue 100
20) Methylene Chloride	2.27	84	16005	8.46	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF070218\
Data File : VF059399.D
Acq On : 2 Jul 2018 20:51
Operator : VA/AP
Sample : J3802-05RE
Misc : 6.01a/5mL/MSVOA-F/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
CF-5(6.0)RE

Quant Time: Jul 03 11:49:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F062518S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 26 08:45:17 2018
Response via : Initial Calibration

