


Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF070518\
Data File : VF059426.D
Acq On : 5 Jul 2018 9:42
Operator : VA/AP
Sample : BFB
Misc : 5.00µ/5mL/MSVOA-F/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F062518S.M
Title : SW846 8260
Last Update : Tue Jun 26 08:45:17 2018



AutoFind: Scans 1110, 1111, 1112; Background Corrected with Scan 1105

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.9	25069	PASS
75	95	30	60	48.8	61560	PASS
95	95	100	100	100.0	126178	PASS
96	95	5	9	7.2	9030	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	79.7	100608	PASS
175	174	5	9	7.4	7446	PASS
176	174	95	101	96.6	97138	PASS
177	176	5	9	6.4	6193	PASS