

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF070618\
 Data File : VF059468.D
 Acq On : 6 Jul 2018 23:48
 Operator : VA/AP
 Sample : J3874-02
 Misc : 5.03g/5mL/MSVOA-F/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 2018-M-NYSEG-MCMaster-PH4-50

Manual Integrations
 APPROVED

apatel
 7/9/2018 12:50:33 PM

Quant Time: Jul 07 04:30:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F062518S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 26 08:45:17 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.01	168	141993	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	5.73	114	217406	50.00	ug/l	-0.03
63) Chlorobenzene-d5	9.89	117	181848	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	12.62	152	85413	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.99	65	113857	62.06	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	124.12%	
35) Dibromofluoromethane	4.26	113	97635	61.46	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	122.92%	
50) Toluene-d8	7.68	98	168137	42.75	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	85.50%	
62) 4-Bromofluorobenzene	11.48	95	77518	34.94	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	69.88%	
Target Compounds						
20) Methylene Chloride	2.31	84	9616m	2.30	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF070618\
Data File : VF059468.D
Acq On : 6 Jul 2018 23:48
Operator : VA/AP
Sample : J3874-02
Misc : 5.03g/5mL/MSVOA-F/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
2018-M-NYSEG-MCMaster-PH4-5(

Manual Integrations
APPROVED

apatel
7/9/2018 12:50:33 PM

Quant Time: Jul 07 04:30:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F062518S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 26 08:45:17 2018
Response via : Initial Calibration

