


Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF071018\
Data File : VF059478.D
Acq On : 10 Jul 2018 10:25
Operator : VA/AP
Sample : BFB
Misc : 5.00µ/5mL/MSVOA F/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F070718S.M
Title : SW846 8260
Last Update : Sat Jul 07 06:35:36 2018



AutoFind: Scans 1111, 1112, 1113; Background Corrected with Scan 1104

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.9	24884	PASS
75	95	30	60	48.6	60674	PASS
95	95	100	100	100.0	124805	PASS
96	95	5	9	6.3	7896	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	85.6	106832	PASS
175	174	5	9	7.7	8195	PASS
176	174	95	101	96.4	102952	PASS
177	176	5	9	6.6	6755	PASS