

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF071818\
 Data File : VF059541.D
 Acq On : 19 Jul 2018 4:34
 Operator : VA/AP
 Sample : J3991-05
 Misc : 4.70g/5mL/MSVOA F/SOIL
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SB-3-B

Manual Integrations
 APPROVED

apatel
 7/19/2018 10:19:04 AM

Quant Time: Jul 19 07:07:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F071718S.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 18 04:09:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.03	168	181668	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	5.76	114	286311	50.00	ug/l	-0.03
63) Chlorobenzene-d5	9.91	117	299781	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	12.63	152	191822	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.00	65	137752	48.76	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	97.52%	
35) Dibromofluoromethane	4.28	113	124963	50.39	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	100.78%	
50) Toluene-d8	7.70	98	259678	41.57	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	83.14%	
62) 4-Bromofluorobenzene	11.50	95	151645	46.28	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	92.56%	
Target Compounds						
20) Methylene Chloride	2.27	84	15018m	7.92	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF071818\
Data File : VF059541.D
Acq On : 19 Jul 2018 4:34
Operator : VA/AP
Sample : J3991-05
Misc : 4.70µ/5mL/MSVOA F/SOIL
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SB-3-B

Manual Integrations
APPROVED

apatel
7/19/2018 10:19:04 AM

Quant Time: Jul 19 07:07:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F071718S.M
Quant Title : SW846 8260
QLast Update : Wed Jul 18 04:09:10 2018
Response via : Initial Calibration

