


Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF072518\
Data File : VF059655.D
Acq On : 25 Jul 2018 12:55
Operator : VA/AP
Sample : BFB
Misc : 5.00µ/5mL/MSVOA F/SOIL
ALS Vial : 80 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F072418S.M
Title : SW846 8260
Last Update : Wed Jul 25 02:29:31 2018



AutoFind: Scans 1113, 1114, 1115; Background Corrected with Scan 1108

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.8	16330	PASS
75	95	30	60	49.8	40957	PASS
95	95	100	100	100.0	82274	PASS
96	95	5	9	6.6	5410	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	88.4	72765	PASS
175	174	5	9	7.9	5764	PASS
176	174	95	101	96.3	70048	PASS
177	176	5	9	7.0	4935	PASS