


Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF073018\
 Data File : VF059685.D
 Acq On : 30 Jul 2018 13:00
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00µ/5mL/MSVOA F/SOIL
 ALS Vial : 80 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F073018S.M
 Title : SW846 8260
 Last Update : Tue Jul 31 05:19:18 2018



AutoFind: Scans 1112, 1113, 1114; Background Corrected with Scan 1107

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.1	21326	PASS
75	95	30	60	51.7	47704	PASS
95	95	100	100	100.0	92202	PASS
96	95	5	9	6.4	5872	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	76.1	70154	PASS
175	174	5	9	8.5	5943	PASS
176	174	95	101	96.3	67560	PASS
177	176	5	9	6.7	4560	PASS