


Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF073118\
Data File : VF059724.D
Acq On : 31 Jul 2018 23:03
Operator : VA/AP
Sample : BFB
Misc : 5.00µ/5mL/MSVOA F/SOIL
ALS Vial : 100 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F073018S.M
Title : SW846 8260
Last Update : Tue Jul 31 15:23:22 2018



AutoFind: Scans 1113, 1114, 1115; Background Corrected with Scan 1107

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.7	12109	PASS
75	95	30	60	55.3	30848	PASS
95	95	100	100	100.0	55762	PASS
96	95	5	9	6.6	3656	PASS
173	174	0.00	2	0.2	101	PASS
174	95	50	100	85.9	47922	PASS
175	174	5	9	7.2	3457	PASS
176	174	95	101	96.4	46189	PASS
177	176	5	9	6.3	2923	PASS