

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_F\DATA\VF080118\
 Data File : VF059780.D
 Acq On : 2 Aug 2018 4:24
 Operator : VA/AP
 Sample : J4164-09
 Misc : 11.74g/5mL/MSVOA_F/SOIL
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SB-2(3-5)

Quant Time: Aug 02 05:36:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F073018S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 31 15:23:22 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	0.00	168	0	0.00	ug/l	-5.03
34) 1,4-Difluorobenzene	5.77	114	60	50.00	ug/l	0.01
63) Chlorobenzene-d5	0.00	117	0	0.00	ug/l	-9.91
72) 1,4-Dichlorobenzene-d4	0.00	152	0	0.00	ug/l	-12.63

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.05	65	75	0.00	ug/l	0.04
Spiked Amount	50.000		Recovery	=	0.00%	
35) Dibromofluoromethane	4.12	113	61	97.88	ug/l	-0.17
Spiked Amount	50.000		Recovery	=	195.76%	
50) Toluene-d8	0.00	98	0	0.00	ug/l	
Spiked Amount	50.000		Recovery	=	0.00%	
62) 4-Bromofluorobenzene	11.56	95	80	99.44	ug/l	0.05
Spiked Amount	50.000		Recovery	=	198.88%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
36) 1,1-Dichloropropene	4.42	75	86	149.445	ug/l #	46
37) Ethyl Acetate	4.31	43	128	466.067	ug/l #	19
38) Carbon Tetrachloride	4.14	117	65	68.137	ug/l #	14
40) Benzene	4.90	78	97	80.054	ug/l #	52
41) Methacrylonitrile	4.95	41	216	1512.195	ug/l #	12
42) 1,2-Dichloroethane	5.06	62	68	80.189	ug/l	85
43) Isopropyl Acetate	6.97	43	102	284.304	ug/l #	46
44) Trichloroethene	5.69	130	66	175.620	ug/l	2
45) 1,2-Dichloropropane	6.39	63	80	250.008	ug/l #	43
48) Methyl methacrylate	6.85	41	163	604.020	ug/l #	13
49) 1,4-Dioxane	6.98	88	63	13317.811	ug/l #	21
51) 4-Methyl-2-Pentanone	8.44	43	77	282.137	ug/l #	42
53) t-1,3-Dichloropropene	8.45	75	65	90.547	ug/l #	41
54) cis-1,3-Dichloropropene	7.55	75	140	201.345	ug/l #	18
56) Ethyl methacrylate	8.74	69	105	306.808	ug/l #	51
57) 1,3-Dichloropropane	8.88	76	159	282.428	ug/l #	41
58) 2-Chloroethyl Vinyl ether	7.79	63	64	1619.694	ug/l	100
59) 2-Hexanone	9.67	43	230	1051.651	ug/l #	30
60) Dibromochloromethane	8.91	129	61	105.343	ug/l #	12


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_F\DATA\VF080118\
Data File : VF059780.D
Acq On : 2 Aug 2018 4:24
Operator : VA/AP
Sample : J4164-09
Misc : 11.74g/5mL/MSVOA_F/SOIL
ALS Vial : 100 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SB-2(3-5)

Quant Time: Aug 02 05:36:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F073018S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 31 15:23:22 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 0.000 ug/l

Expected RT: 5.03 min

Lab File: VF059780.D

Acq: 2 Aug 2018 4:24

Tgt Ion: 168

Sig Exp Ratio

168 100



99 79.1

Instrument :

MSVOA_F

ClientSampleId :

SB-2(3-5)



#33

1,2-Dichloroethane-d4

Concen: 0.000 ug/l

RT: 5.05 min Scan# 459

Delta R.T. 0.04 min

Lab File: VF059780.D




Acq: 2 Aug 2018 4:24


Tgt Ion: 65 Resp: 75

Ion Ratio Lower Upper

65 100

67 0.0 0.0 84.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.77 min Scan# 532

Delta R.T. 0.01 min

Lab File: VF059780.D

Acq: 2 Aug 2018 4:24

Instrument :

MSVOA_F

ClientSampleId :

SB-2(3-5)


Tgt Ion:114 Resp: 60

Ion Ratio Lower Upper

114 100

63 0.0 0.0 47.4

88 0.0 0.0 41.2




Abundance Ion 114.00 (113.70 to 114.70): V


Ion 63.00 (62.70 to 63.70): VF0

Ion 88.00 (87.70 to 88.70): VF0

Time-->



Time-->



#35

Dibromofluoromethane

Concen: 97.883 ug/l

RT: 4.12 min Scan# 364

Delta R.T. -0.17 min

Lab File: VF059780.D

Acq: 2 Aug 2018 4:24


Tgt Ion:113 Resp: 61

Ion Ratio Lower Upper

113 100

111 0.0 79.0 118.4#

192 0.0 13.2 19.8#




Abundance Ion 113.00 (112.70 to 113.70): V


Ion 111.00 (110.70 to 111.70): V

Ion 192.00 (191.70 to 192.70): V

Time-->







Time-->



#36
 1,1-Dichloropropene
 Concen: 149.445 ug/l
 RT: 4.42 min Scan# 395
 Delta R.T. -0.03 min
 Lab File: VF059780.D
 Acq: 2 Aug 2018 4:24




Instrument :
 MSVOA_F
 ClientSampleId :
 SB-2(3-5)


Tgt Ion: 75 Resp: 86
 Ion Ratio Lower Upper
 75 100
 110 0.0 14.0 41.9#
 77 0.0 24.2 36.2#



#37
 Ethyl Acetate
 Concen: 466.067 ug/l
 RT: 4.31 min Scan# 383
 Delta R.T. 0.01 min
 Lab File: VF059780.D
 Acq: 2 Aug 2018 4:24

Tgt Ion: 43 Resp: 128
 Ion Ratio Lower Upper
 43 100
 61 0.0 58.2 87.2#
 70 0.0 6.0 9.0#





#38

Carbon Tetrachloride

Concen: 68.137 ug/l

RT: 4.14 min Scan# 366

Delta R.T. -0.03 min

Lab File: VF059780.D

Acq: 2 Aug 2018 4:24

Instrument :

MSVOA_F

ClientSampled :

SB-2(3-5)


Tgt Ion: 117 Resp: 65

Ion Ratio Lower Upper

117 100

119 0.0 72.9 109.3#




121 0.0 25.0 37.4#



Abundance Ion 117.00 (116.70 to 117.70): V

Ion 119.00 (118.70 to 119.70): V

Ion 121.00 (120.70 to 121.70): V



#40

Benzene

Concen: 80.054 ug/l

RT: 4.90 min Scan# 443

Delta R.T. 0.09 min

Lab File: VF059780.D


Acq: 2 Aug 2018 4:24

Tgt Ion: 78 Resp: 97

Ion Ratio Lower Upper



78 100


77 0.0 18.9 28.3#



Abundance Ion 78.10 (77.80 to 78.80): VF0

Ion 77.00 (76.70 to 77.70): VF0





#41

Methacrylonitrile

Concen: 1512.195 ug/l

RT: 4.95 min Scan# 448

Delta R.T. 0.03 min

Lab File: VF059780.D

Acq: 2 Aug 2018 4:24

Instrument :

MSVOA_F

ClientSampleId :

SB-2(3-5)

Tgt Ion: 41 Resp: 216





Ion Ratio Lower Upper

41 100

39 156.3 55.2 82.8#

67 0.0 36.2 54.2#

52 99.3 37.7 56.5#



#42

1,2-Dichloroethane

Concen: 80.189 ug/l

RT: 5.06 min Scan# 460

Delta R.T. -0.05 min

Lab File: VF059780.D




Acq: 2 Aug 2018 4:24


Tgt Ion: 62 Resp: 68

Ion Ratio Lower Upper

62 100

98 0.0 0.0 9.6





#43

Isopropyl Acetate

Concen: 284.304 ug/l

RT: 6.97 min Scan# 653

Delta R.T. -0.15 min

Lab File: VF059780.D

Acq: 2 Aug 2018 4:24

Instrument :

MSVOA_F

ClientSampleId :

SB-2(3-5)


Tgt Ion: 43 Resp: 102

Ion Ratio Lower Upper

43 100

61 0.0 23.8 35.8#




87 0.0 0.6 1.0#



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 87.00 (86.70 to 87.70): VF0



#44

Trichloroethene

Concen: 175.620 ug/l

RT: 5.69 min Scan# 523

Delta R.T. 0.01 min

Lab File: VF059780.D


Acq: 2 Aug 2018 4:24

Tgt Ion: 130 Resp: 66

Ion Ratio Lower Upper



130 100


95 0.0 0.0 191.0



Abundance Ion 129.90 (129.60 to 130.60): V

Ion 94.90 (94.60 to 95.60): VF0





#45

1,2-Dichloropropane

Concen: 250.008 ug/l

RT: 6.39 min Scan# 595

Delta R.T. 0.00 min

Lab File: VF059780.D

Acq: 2 Aug 2018 4:24

Instrument :

MSVOA_F

ClientSampleId :


SB-2(3-5)

Tgt Ion: 63 Resp: 80

Ion Ratio Lower Upper

63 100

65 0.0 25.2 37.8#



Abundance Ion 63.00 (62.70 to 63.70): VF0

Ion 65.00 (64.70 to 65.70): VF0

6.39

150



100

50

0

6.36 6.38 6.40 6.42

Time-->



#48

Methyl methacrylate

Concen: 604.020 ug/l

RT: 6.85 min Scan# 641

Delta R.T. -0.02 min

Lab File: VF059780.D

Acq: 2 Aug 2018 4:24


Tgt Ion: 41 Resp: 163

Ion Ratio Lower Upper

41 100

69 0.0 44.4 66.6#

39 0.0 69.4 104.2#



Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 69.00 (68.70 to 69.70): VF0

Ion 39.00 (38.70 to 39.70): VF0

200

150


100


50

0

6.80 6.82 6.84 6.86 6.88

Time-->





#49

1,4-Dioxane

Concen: 13317.811 ug/l

RT: 6.98 min Scan# 654

Delta R.T. 0.12 min

Lab File: VF059780.D

Acq: 2 Aug 2018 4:24

Instrument :


MSVOA_F

ClientSampleId :

SB-2(3-5)

Tgt Ion: 88 Resp: 63




Ion	Ratio	Lower	Upper
88	100		
43	0.0	45.2	67.8#
58	0.0	49.0	73.6#



Abundance Ion 88.00 (87.70 to 88.70): VF0

Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0



#50

Toluene-d8

Concen: 0.000 ug/l

Expected RT: 7.71 min

Lab File: VF059780.D


Acq: 2 Aug 2018 4:24


Tgt Ion: 98

Sig Exp Ratio

98 100

100 66.0









#51
 4-Methyl-2-Pentanone
 Concen: 282.137 ug/l
 RT: 8.44 min Scan# 802
 Delta R.T. 0.03 min
 Lab File: VF059780.D
 Acq: 2 Aug 2018 4:24

Instrument :
 MSVOA_F
 ClientSampleId :
 SB-2(3-5)

Tgt Ion: 43 Resp: 77
 Ion Ratio Lower Upper
 43 100
 58 0.0 25.9 38.9#




Abundance Ion 43.00 (42.70 to 43.70): VF0
 Ion 58.00 (57.70 to 58.70): VF0






#53
 t-1,3-Dichloropropene
 Concen: 90.547 ug/l
 RT: 8.45 min Scan# 804
 Delta R.T. 0.03 min
 Lab File: VF059780.D
 Acq: 2 Aug 2018 4:24

Tgt Ion: 75 Resp: 65
 Ion Ratio Lower Upper
 75 100
 77 0.0 27.0 40.4#



Abundance Ion 75.00 (74.70 to 75.70): VF0
 Ion 77.00 (76.70 to 77.70): VF0





#54

cis-1,3-Dichloropropene

Concen: 201.345 ug/l

RT: 7.55 min Scan# 712

Delta R.T. 0.10 min

Lab File: VF059780.D

Acq: 2 Aug 2018 4:24

Instrument :

MSVOA_F

ClientSampled :

SB-2(3-5)





Tgt Ion: 75 Resp: 140

Ion Ratio Lower Upper

75 100

77 0.0 24.0 36.0#

39 0.0 65.4 98.2#



#56

Ethyl methacrylate

Concen: 306.808 ug/l

RT: 8.74 min Scan# 833

Delta R.T. -0.02 min

Lab File: VF059780.D

Acq: 2 Aug 2018 4:24




Tgt Ion: 69 Resp: 105


Ion Ratio Lower Upper

69 100

41 95.5 62.3 93.5#

39 0.0 52.4 78.6#





#57

1,3-Dichloropropane

Concen: 282.428 ug/l

RT: 8.88 min Scan# 847

Delta R.T. -0.12 min

Lab File: VF059780.D

Acq: 2 Aug 2018 4:24

Instrument :

MSVOA_F

ClientSampleId :


SB-2(3-5)

Tgt Ion: 76 Resp: 159

Ion Ratio Lower Upper

76 100

78 0.0 27.2 40.8#



Abundance Ion 76.00 (75.70 to 76.70): VF0

Ion 78.00 (77.70 to 78.70): VF0

8.88



150

100

50

0

Time--> 8.84 8.86 8.88 8.90 8.92



#58

2-Chloroethyl Vinyl ether

Concen: 1619.694 ug/l

RT: 7.79 min Scan# 737

Delta R.T. 0.02 min

Lab File: VF059780.D


Acq: 2 Aug 2018 4:24

Tgt Ion: 63 Resp: 64

Ion Ratio Lower Upper

63 100

106 0.0 0.0 0.0



Abundance Ion 63.00 (62.70 to 63.70): VF0

Ion 106.00 (105.70 to 106.70): V


150


100

50

0

Time--> 7.76 7.78 7.80 7.82









#59
2-Hexanone
Concen: 1051.651 ug/l
RT: 9.67 min Scan# 927
Delta R.T. 0.04 min
Lab File: VF059780.D
Acq: 2 Aug 2018 4:24




Instrument :
MSVOA_F
ClientSampleId :
SB-2(3-5)


Tgt Ion: 43 Resp: 230
Ion Ratio Lower Upper
43 100
58 0.0 23.3 69.8#



#60
Dibromochloromethane
Concen: 105.343 ug/l
RT: 8.91 min Scan# 850
Delta R.T. 0.05 min
Lab File: VF059780.D
Acq: 2 Aug 2018 4:24

Tgt Ion: 129 Resp: 61
Ion Ratio Lower Upper
129 100
127 0.0 37.7 113.1#





#62

4-Bromofluorobenzene

Concen: 99.437 ug/l

RT: 11.56 min Scan# 1119

Delta R.T. 0.05 min

Lab File: VF059780.D

Acq: 2 Aug 2018 4:24

Instrument :

MSVOA_F

ClientSampleId :

SB-2(3-5)


Tgt Ion: 95 Resp: 80

Ion Ratio Lower Upper

95 100

174 0.0 0.0 178.4




176 0.0 0.0 173.0



Abundance Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V



#63

Chlorobenzene-d5

Concen: 0.000 ug/l

Expected RT: 9.91 min

Lab File: VF059780.D

Acq: 2 Aug 2018 4:24


Tgt Ion: 117


Sig Exp Ratio

117 100

82 61.2

119 34.1





#72
 1,4-Dichlorobenzene-d4
 Concen: 0.000 ug/l
 Expected RT: 12.63 min
 Lab File: VF059780.D
 Acq: 2 Aug 2018 4:24

Instrument :
 MSVOA_F
 ClientSampleId :
 SB-2(3-5)

Tgt Ion: 152
 Sig Exp Ratio
 152 100
 115 56.4
 150 158.1

