

Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF082418\
 Data File : VF060099.D
 Acq On : 24 Aug 2018 14:52
 Operator : VA/AP
 Sample : J4581-02
 Misc : 5.74g/5mL/MSVOA F/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 EP1-Q3-A2

Quant Time: Sep 04 14:32:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F082118S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 04:43:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.04	168	182720	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.76	114	276066	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.91	117	311403	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	12.63	152	183074	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.01	65	120276	48.10	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	96.20%	
35) Dibromofluoromethane	4.28	113	113697	52.45	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	104.90%	
50) Toluene-d8	7.71	98	284169	46.16	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	92.32%	
62) 4-Bromofluorobenzene	11.51	95	157690	48.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.84%	
Target Compounds						
59) 2-Hexanone	9.62	43	5560	4.07	ug/l	80
64) Tetrachloroethene	8.31	164	5185	2.47	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF082418\
Data File : VF060099.D
Acq On : 24 Aug 2018 14:52
Operator : VA/AP
Sample : J4581-02
Misc : 5.74g/5mL/MSVOA F/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
EP1-Q3-A2

Quant Time: Sep 04 14:32:20 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F082118S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 04:43:43 2018
Response via : Initial Calibration

