

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF082418\
 Data File : VF060102.D
 Acq On : 24 Aug 2018 16:16
 Operator : VA/AP
 Sample : J4581-19
 Misc : 5.53g/5mL/MSVOA F/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 EP1-Q3-D3

Quant Time: Aug 25 04:49:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F082118S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 04:43:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.03	168	160739	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	5.76	114	233378	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.91	117	261437	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	12.63	152	146748	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.00	65	106642	48.48	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	96.96%	
35) Dibromofluoromethane	4.28	113	106776	58.26	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	116.52%	
50) Toluene-d8	7.70	98	271595	52.18	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	104.36%	
62) 4-Bromofluorobenzene	11.51	95	134715	48.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.86%	
Target Compounds						
20) Methylene Chloride	2.28	84	9965	4.04	ug/l	# 69
30) Chloroform	4.01	83	3706	1.05	ug/l	92
64) Tetrachloroethene	8.30	164	17748	10.08	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF082418\
Data File : VF060102.D
Acq On : 24 Aug 2018 16:16
Operator : VA/AP
Sample : J4581-19
Misc : 5.53g/5mL/MSVOA F/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
EP1-Q3-D3

Quant Time: Aug 25 04:49:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F082118S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 04:43:43 2018
Response via : Initial Calibration

