

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF082418\
 Data File : VF060114.D
 Acq On : 24 Aug 2018 21:50
 Operator : VA/AP
 Sample : J4636-01
 Misc : 6.02g/5mL/MSVOA F/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 BYR-2-E(0-1)

Quant Time: Aug 25 05:24:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F082118S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 22 04:43:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.03	168	128061	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	5.75	114	195851	50.00	ug/l	-0.03
63) Chlorobenzene-d5	9.90	117	153120	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	12.62	152	41599	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.00	65	66133	37.73	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	75.46%	
35) Dibromofluoromethane	4.27	113	70304	45.71	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	91.42%	
50) Toluene-d8	7.70	98	142351	32.59	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	65.18%	
62) 4-Bromofluorobenzene	11.50	95	42009	18.18	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	36.36%	
Target Compounds						
16) Acetone	2.32	43	4834	10.49	ug/l #	66
20) Methylene Chloride	2.28	84	8716	4.99	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF082418\
Data File : VF060114.D
Acq On : 24 Aug 2018 21:50
Operator : VA/AP
Sample : J4636-01
Misc : 6.02g/5mL/MSVOA F/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
BYR-2-E(0-1)

Quant Time: Aug 25 05:24:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F082118S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 22 04:43:43 2018
Response via : Initial Calibration

