

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF082918\
 Data File : VF060183.D
 Acq On : 29 Aug 2018 8:44
 Operator : VA/AP
 Sample : J4277-09
 Misc : 6.48g/5mL/MSVOA F/SOIL
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 OR-01-082718-A

Quant Time: Aug 29 10:58:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F082918S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 29 03:22:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.04	168	55111	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	5.76	114	64149	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.91	117	26616	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.63	152	8400	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	5.01	65	15985	46.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.88%	
35) Dibromofluoromethane	4.28	113	20055	48.60	ug/l	0.01
Spiked Amount 50.000			Recovery	=	97.20%	
50) Toluene-d8	7.71	98	28445	20.15	ug/l	0.01
Spiked Amount 50.000			Recovery	=	40.30%	
62) 4-Bromofluorobenzene	11.50	95	12145	23.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	46.66%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	2.29	84	1921	3.58	ug/l	# 56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF082918\
Data File : VF060183.D
Acq On : 29 Aug 2018 8:44
Operator : VA/AP
Sample : J4277-09
Misc : 6.48g/5mL/MSVOA F/SOIL
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
OR-01-082718-A

Quant Time: Aug 29 10:58:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F082918S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 29 03:22:03 2018
Response via : Initial Calibration

