

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF090718\
 Data File : VF060220.D
 Acq On : 7 Sep 2018 19:45
 Operator : VA/AP
 Sample : J4803-03
 Misc : 5.27g/5mL/MSVOA F/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 EP1-Q2-E7

Quant Time: Sep 08 00:28:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F090618S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 07 10:44:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.04	168	158662	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.77	114	230743	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.92	117	243800	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.63	152	133982	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.01	65	116463	42.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.04%	
35) Dibromofluoromethane	4.28	113	101029	44.69	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	89.38%	
50) Toluene-d8	7.71	98	208203	36.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.82%	
62) 4-Bromofluorobenzene	11.51	95	107786	33.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.72%	
Target Compounds						
20) Methylene Chloride	2.28	84	15799	3.37	ug/l	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF090718\
Data File : VF060220.D
Acq On : 7 Sep 2018 19:45
Operator : VA/AP
Sample : J4803-03
Misc : 5.27g/5mL/MSVOA F/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
EP1-Q2-E7

Quant Time: Sep 08 00:28:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F090618S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 07 10:44:05 2018
Response via : Initial Calibration

