

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF091018\
 Data File : VF060240.D
 Acq On : 10 Sep 2018 17:42
 Operator : VA/AP
 Sample : J4803-16
 Misc : 5.23/5mL/MSVOA F/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 EP1-Q2-J5

Quant Time: Sep 11 08:10:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F090618S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 07 10:44:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.05	168	207461	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	5.77	114	305217	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.92	117	304346	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.63	152	179234	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.02	65	107952	29.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	59.58%	
35) Dibromofluoromethane	4.29	113	102798	34.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.76%	
50) Toluene-d8	7.71	98	265689	35.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.22%	
62) 4-Bromofluorobenzene	11.52	95	139076	33.03	ug/l	0.01
Spiked Amount 50.000			Recovery	=	66.06%	
Target Compounds						
80) 1,3,5-Trimethylbenzene	11.91	105	13076	1.20	ug/l	99
84) 1,2,4-Trimethylbenzene	12.29	105	24670	2.23	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF091018\
Data File : VF060240.D
Acq On : 10 Sep 2018 17:42
Operator : VA/AP
Sample : J4803-16
Misc : 5.23/5mL/MSVOA F/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
EP1-Q2-J5

Quant Time: Sep 11 08:10:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F090618S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 07 10:44:05 2018
Response via : Initial Calibration

