

Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF092518\
 Data File : VF060324.D
 Acq On : 25 Sep 2018 13:58
 Operator : VA/AP
 Sample : VSTDCCC050
 Misc : 5.00/5mL/MSVOA F/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_F
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 26 02:08:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F091918S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 20 02:58:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	-0.02
2 T	Dichlorodifluoromethane	0.888	0.782	11.9	106	0.00
3 P	Chloromethane	0.681	0.595	12.6	101	0.00
4 C	Vinyl Chloride	0.694	0.614	11.5#	102	0.00
5 T	Bromomethane	0.466	0.443	4.9	107	0.00
6 T	Chloroethane	0.331	0.342	-3.3	112	0.00
7 T	Trichlorofluoromethane	1.294	1.372	-6.0	113	0.00
8 T	Diethyl Ether	0.215	0.204	5.1	108	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.688	0.683	0.7	109	0.00
10 T	Methyl Iodide	1.109	1.073	3.2	108	0.00
11 T	Tert butyl alcohol	0.043	0.044	-2.3	125	-0.02
12 CM	1,1-Dichloroethene	0.569	0.574	-0.9#	108	0.00
13 T	Acrolein	0.021	0.014	33.3#	127	0.00
14 T	Allyl chloride	0.993	0.938	5.5	101	0.00
15 T	Acrylonitrile	0.095	0.102	-7.4	120	0.00
16 T	Acetone	0.190	0.206	-8.4	115	0.00
17 T	Carbon Disulfide	1.717	1.497	12.8	103	0.00
18 T	Methyl Acetate	0.316	0.318	-0.6	110	0.00
19 T	Methyl tert-butyl Ether	1.418	1.412	0.4	109	0.00
20 T	Methylene Chloride	0.665	0.566	14.9	112	0.00
21 T	trans-1,2-Dichloroethene	0.588	0.557	5.3	108	0.00
22 T	Diisopropyl ether	1.873	1.821	2.8	106	0.00
23 T	Vinyl Acetate	0.736	0.686	6.8	103	0.00
24 P	1,1-Dichloroethane	1.215	1.230	-1.2	117	0.00
25 T	2-Butanone	0.255	0.238	6.7	96	0.00
26 T	2,2-Dichloropropane	0.693	0.663	4.3	100	-0.02
27 T	cis-1,2-Dichloroethene	0.772	0.683	11.5	98	-0.02
28 T	Bromochloromethane	0.473	0.433	8.5	102	-0.02
29	Tetrahydrofuran	0.093	0.079	15.1	93	0.00
30 C	Chloroform	1.466	1.406	4.1#	100	-0.02
31 T	Cyclohexane	1.026	0.850	17.2	94	-0.02
32 T	1,1,1-Trichloroethane	1.157	1.098	5.1	99	-0.02
33 S	1,2-Dichloroethane-d4	0.673	0.622	7.6	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
35 S	Dibromofluoromethane	0.446	0.386	13.5	94	-0.02
36 T	1,1-Dichloropropene	0.624	0.553	11.4	99	-0.02
37 T	Ethyl Acetate	0.264	0.206	22.0	88	-0.02
38 T	Carbon Tetrachloride	0.670	0.587	12.4	97	-0.02
39 T	Methylcyclohexane	0.676	0.600	11.2	98	0.00
40 TM	Benzene	1.414	1.212	14.3	96	0.00
41 T	Methacrylonitrile	0.143	0.117	18.2	90	-0.02
42 TM	1,2-Dichloroethane	0.521	0.499	4.2	107	0.00
43 T	Isopropyl Acetate	0.318	0.270	15.1	98	0.00
44 TM	Trichloroethene	0.421	0.393	6.7	102	0.00
45 C	1,2-Dichloropropane	0.366	0.337	7.9#	102	-0.02

Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF092518\
 Data File : VF060324.D
 Acq On : 25 Sep 2018 13:58
 Operator : VA/AP
 Sample : VSTDCCC050
 Misc : 5.00/5mL/MSVOA F/SOIL
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_F
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 26 02:08:30 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F091918S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 20 02:58:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.262	0.243	7.3	103	-0.02
47 T	Bromodichloromethane	0.629	0.618	1.7	109	-0.02
48 T	Methyl methacrylate	0.223	0.195	12.6	103	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	100	-0.02
50 S	Toluene-d8	1.161	1.026	11.6	93	-0.02
51 T	4-Methyl-2-Pentanone	0.245	0.266	-8.6	114	-0.02
52 CM	Toluene	0.895	0.811	9.4#	102	-0.02
53 T	t-1,3-Dichloropropene	0.507	0.473	6.7	104	0.00
54 T	cis-1,3-Dichloropropene	0.639	0.560	12.4	95	0.00
55 T	1,1,2-Trichloroethane	0.272	0.260	4.4	101	-0.02
56 T	Ethyl methacrylate	0.328	0.327	0.3	113	0.00
57 T	1,3-Dichloropropane	0.450	0.427	5.1	101	0.00
58 T	2-Chloroethyl Vinyl ether	0.029	0.027	6.9	106	0.00
59 T	2-Hexanone	0.174	0.193	-10.9	115	0.00
60 T	Dibromochloromethane	0.405	0.386	4.7	101	-0.02
61 T	1,2-Dibromoethane	0.302	0.281	7.0	104	0.00
62 S	4-Bromofluorobenzene	0.502	0.451	10.2	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	104	-0.02
64 T	Tetrachloroethene	0.453	0.404	10.8	97	0.00
65 PM	Chlorobenzene	1.133	1.075	5.1	102	0.00
66 T	1,1,1,2-Tetrachloroethane	0.470	0.459	2.3	108	0.00
67 C	Ethyl Benzene	2.095	1.966	6.2#	102	0.00
68 T	m/p-Xylenes	0.724	0.680	6.1	104	0.00
69 T	o-Xylene	0.812	0.804	1.0	107	0.00
70 T	Styrene	1.145	1.114	2.7	106	-0.02
71 P	Bromoform	0.258	0.248	3.9	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
73 T	Isopropylbenzene	4.489	4.169	7.1	106	0.00
74 T	N-amyl acetate	0.957	0.933	2.5	117	0.00
75 P	1,1,2,2-Tetrachloroethane	0.795	0.770	3.1	108	0.00
76 T	1,2,3-Trichloropropane	0.570	0.527	7.5	107	-0.02
77 T	Bromobenzene	0.967	0.873	9.7	102	0.00
78 T	n-propylbenzene	5.299	4.948	6.6	106	0.00
79 T	2-Chlorotoluene	3.028	2.796	7.7	109	0.00
80 T	1,3,5-Trimethylbenzene	3.613	3.512	2.8	107	0.00
81 T	trans-1,4-Dichloro-2-butene	0.301	0.278	7.6	111	0.00
82 T	4-Chlorotoluene	3.043	2.869	5.7	108	0.00
83 T	tert-Butylbenzene	3.663	3.483	4.9	107	0.00
84 T	1,2,4-Trimethylbenzene	3.669	3.425	6.7	109	0.00
85 T	sec-Butylbenzene	5.071	4.817	5.0	106	0.00
86 T	p-Isopropyltoluene	4.013	3.746	6.7	104	0.00
87 T	1,3-Dichlorobenzene	1.808	1.614	10.7	103	0.00
88 T	1,4-Dichlorobenzene	1.787	1.636	8.4	102	0.00
89 T	n-Butylbenzene	4.245	3.954	6.9	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA F\Data\VF092518\
Data File : VF060324.D
Acq On : 25 Sep 2018 13:58
Operator : VA/AP
Sample : VSTDCCC050
Misc : 5.00/5mL/MSVOA F/SOIL
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_F
LabSampleId :
VSTDCCC050

Quant Time: Sep 26 02:08:30 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F091918S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 20 02:58:42 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	1.025	1.023	0.2	110	0.00
91 T	1,2-Dichlorobenzene	1.724	1.628	5.6	109	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.149	0.139	6.7	108	0.00
93 T	1,2,4-Trichlorobenzene	1.253	1.161	7.3	107	0.00
94 T	Hexachlorobutadiene	0.847	0.826	2.5	108	0.00
95 T	Naphthalene	2.197	2.271	-3.4	119	0.00
96 T	1,2,3-Trichlorobenzene	1.158	1.166	-0.7	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6