

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF100218\
 Data File : VF060413.D
 Acq On : 2 Oct 2018 20:19
 Operator : VA/AP
 Sample : J5150-02
 Misc : 5.19/5mL/MSVOA F/SOIL
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SSSI-0926-S-10B

Manual Integrations
 APPROVED

apatel
 10/3/2018 1:15:50 PM

Quant Time: Oct 03 03:25:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F100118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 11:42:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.02	168	250279	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.75	114	430559	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.90	117	330512	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.63	152	146058	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.00	65	207048	52.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.74%	
35) Dibromofluoromethane	4.27	113	223175	55.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.36%	
50) Toluene-d8	7.69	98	435374	43.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.42%	
62) 4-Bromofluorobenzene	11.49	95	168944	37.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.28%	
Target Compounds						
16) Acetone	2.33	43	1572827	2127.802	ug/l #	88
20) Methylene Chloride	2.32	84	25945m	3.947	ug/l	
95) Naphthalene	14.51	128	113665	18.351	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF100218\
Data File : VF060413.D
Acq On : 2 Oct 2018 20:19
Operator : VA/AP
Sample : J5150-02
Misc : 5.19/5mL/MSVOA F/SOIL
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SSSI-0926-S-10B

Manual Integrations
APPROVED

apatel
10/3/2018 1:15:50 PM

Quant Time: Oct 03 03:25:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F100118S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 11:42:56 2018
Response via : Initial Calibration

