

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF100318\
 Data File : VF060424.D
 Acq On : 3 Oct 2018 19:40
 Operator : VA/AP
 Sample : J5223-01
 Misc : 5.77/5mL/MSVOA F/SOIL
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SSSI-1001-S-DH-01

Manual Integrations
 APPROVED

apatel
 10/4/2018 2:06:47 PM

Quant Time: Oct 04 06:02:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F100118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 11:42:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.02	168	222277	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	5.75	114	372077	50.00	ug/l	0.01
63) Chlorobenzene-d5	9.90	117	332164	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	12.62	152	198263	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.00	65	181081	52.06	ug/l	0.01
Spiked Amount 50.000			Recovery	=	104.12%	
35) Dibromofluoromethane	4.27	113	179031	51.22	ug/l	0.01
Spiked Amount 50.000			Recovery	=	102.44%	
50) Toluene-d8	7.69	98	404083	46.94	ug/l	0.01
Spiked Amount 50.000			Recovery	=	93.88%	
62) 4-Bromofluorobenzene	11.50	95	200256	50.95	ug/l	0.01
Spiked Amount 50.000			Recovery	=	101.90%	
Target Compounds						
16) Acetone	2.33	43	47928m	47.722	ug/l	Qvalue
20) Methylene Chloride	2.30	84	22296m	3.617	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF100318\
Data File : VF060424.D
Acq On : 3 Oct 2018 19:40
Operator : VA/AP
Sample : J5223-01
Misc : 5.77/5mL/MSVOA F/SOIL
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SSSI-1001-S-DH-01

Manual Integrations
APPROVED

apatel
10/4/2018 2:06:47 PM

Quant Time: Oct 04 06:02:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F100118S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 11:42:56 2018
Response via : Initial Calibration

