

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF100318\
 Data File : VF060426.D
 Acq On : 3 Oct 2018 20:35
 Operator : VA/AP
 Sample : J5192-03
 Misc : 5.98/5mL/MSVOA F/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 AU-05-100218-B

Quant Time: Oct 04 06:39:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F100118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 11:42:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.01	168	259553	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.74	114	436142	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.90	117	342342	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	12.62	152	147335	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.99	65	147113	36.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	72.44%	
35) Dibromofluoromethane	4.26	113	150540	36.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.48%	
50) Toluene-d8	7.68	98	287464	28.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	56.98%	
62) 4-Bromofluorobenzene	11.50	95	124460	27.01	ug/l	0.01
Spiked Amount 50.000			Recovery	=	54.02%	
Target Compounds						
20) Methylene Chloride	2.27	84	33630	6.500	ug/l	88
95) Naphthalene	14.51	128	14262	2.283	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF100318\
Data File : VF060426.D
Acq On : 3 Oct 2018 20:35
Operator : VA/AP
Sample : J5192-03
Misc : 5.98/5mL/MSVOA F/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
AU-05-100218-B

Quant Time: Oct 04 06:39:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F100118S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 11:42:56 2018
Response via : Initial Calibration

