

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF100418\
 Data File : VF060446.D
 Acq On : 4 Oct 2018 20:01
 Operator : VA/AP
 Sample : J5218-01
 Misc : 5.26/5mL/MSVOA F/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 280-356

Quant Time: Oct 05 14:28:52 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F100118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 11:42:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.03	168	146934	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	5.76	114	251678	50.00	ug/l	0.02
63) Chlorobenzene-d5	9.91	117	152711	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	12.63	152	51320	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.01	65	148821	64.73	ug/l	0.02
Spiked Amount 50.000			Recovery	=	129.46%	
35) Dibromofluoromethane	4.28	113	140429	59.39	ug/l	0.02
Spiked Amount 50.000			Recovery	=	118.78%	
50) Toluene-d8	7.70	98	138441	23.78	ug/l	0.02
Spiked Amount 50.000			Recovery	=	47.56%	
62) 4-Bromofluorobenzene	11.50	95	81424	30.63	ug/l	0.02
Spiked Amount 50.000			Recovery	=	61.26%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.31	84	11029	1.129	ug/l	88
40) Benzene	4.79	78	12739	2.068	ug/l	99
73) Isopropylbenzene	11.22	105	8045	1.991	ug/l	90
78) n-propylbenzene	11.68	91	5612	1.146	ug/l	83
83) tert-Butylbenzene	12.20	119	238874	70.544	ug/l	85
84) 1,2,4-Trimethylbenzene	12.28	105	4670	1.372	ug/l	93
85) sec-Butylbenzene	12.38	105	9864	2.110	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF100418\
Data File : VF060446.D
Acq On : 4 Oct 2018 20:01
Operator : VA/AP
Sample : J5218-01
Misc : 5.26/5mL/MSVOA F/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
280-356

Quant Time: Oct 05 14:28:52 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F100118S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 11:42:56 2018
Response via : Initial Calibration

