

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF100418\
 Data File : VF060448.D
 Acq On : 4 Oct 2018 20:55
 Operator : VA/AP
 Sample : J5218-05
 Misc : 3.02/5mL/MSVOA F/SOIL
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 293-230

Manual Integrations
 APPROVED

apatel
 10/5/2018 4:58:37 PM

Quant Time: Oct 05 15:06:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F100118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 11:42:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.03	168	164049	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	5.75	114	242477	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.91	117	135104	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	12.63	152	31531	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.01	65	134791	52.51	ug/l	0.02
Spiked Amount 50.000			Recovery	=	105.02%	
35) Dibromofluoromethane	4.28	113	27549	12.09	ug/l	0.02
Spiked Amount 50.000			Recovery	=	24.18%	
50) Toluene-d8	7.70	98	193013	34.40	ug/l	0.02
Spiked Amount 50.000			Recovery	=	68.80%	
62) 4-Bromofluorobenzene	11.50	95	51740	20.20	ug/l	0.02
Spiked Amount 50.000			Recovery	=	40.40%	
Target Compounds						
16) Acetone	2.34	43	22874	21.608	ug/l #	88
20) Methylene Chloride	2.27	84	12786m	1.413	ug/l	
27) cis-1,2-Dichloroethene	3.63	96	13460	6.058	ug/l	78
40) Benzene	4.80	78	64828	10.923	ug/l	97
44) Trichloroethene	5.66	130	88007	46.216	ug/l	95
52) Toluene	7.77	92	20827	5.167	ug/l	95
68) m/p-Xylenes	10.25	106	5093	2.808	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF100418\
Data File : VF060448.D
Acq On : 4 Oct 2018 20:55
Operator : VA/AP
Sample : J5218-05
Misc : 3.02/5mL/MSVOA F/SOIL
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
293-230

Manual Integrations
APPROVED

apatel
10/5/2018 4:58:37 PM

Quant Time: Oct 05 15:06:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F100118S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 11:42:56 2018
Response via : Initial Calibration

