

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_F\DATA\VF100818\
 Data File : VF060487.D
 Acq On : 8 Oct 2018 16:32
 Operator : VA/AP
 Sample : J5084-20RE
 Misc : 6.76/5mL/MSVOA_F/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 B-3(14.5-15)RE

Manual Integrations
 APPROVED

apatel
 10/9/2018 8:53:41 AM

Quant Time: Oct 09 02:16:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F100118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 11:42:56 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.01	168	346040	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.73	114	569159	50.00	ug/l	-0.01
63) Chlorobenzene-d5	9.89	117	485630	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.61	152	275004	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.98	65	170415	31.47	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	62.94%	
35) Dibromofluoromethane	4.26	113	196434	36.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.48%	
50) Toluene-d8	7.67	98	426070	32.36	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	64.72%	
62) 4-Bromofluorobenzene	11.48	95	203678	33.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.76%	
Target Compounds						
16) Acetone	2.32	43	42617	16.028	ug/l #	92
20) Methylene Chloride	2.32	84	26474m	1.271	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_F\DATA\VF100818\
Data File : VF060487.D
Acq On : 8 Oct 2018 16:32
Operator : VA/AP
Sample : J5084-20RE
Misc : 6.76/5mL/MSVOA_F/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
B-3(14.5-15)RE

Manual Integrations
APPROVED

apatel
10/9/2018 8:53:41 AM

Quant Time: Oct 09 02:16:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F100118S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 11:42:56 2018
Response via : Initial Calibration

