

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF102618\
 Data File : VF060516.D
 Acq On : 26 Oct 2018 15:01
 Operator : VA/AP
 Sample : VF1026SBL01
 Misc : 5.00/5mL/MSVOA F/SOIL
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VF1026SBL01

Quant Time: Oct 27 02:12:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F102518S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 26 02:17:41 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.03	168	227331	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	5.75	114	360771	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.91	117	309419	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.63	152	195441	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.00	65	206671	45.74	ug/l	-0.02
Spiked Amount	50.000		Recovery	=	91.48%	
35) Dibromofluoromethane	4.27	113	208656	48.25	ug/l	-0.02
Spiked Amount	50.000		Recovery	=	96.50%	
50) Toluene-d8	7.70	98	379840	43.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.54%	
62) 4-Bromofluorobenzene	11.51	95	191209	39.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.12%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF102618\
Data File : VF060516.D
Acq On : 26 Oct 2018 15:01
Operator : VA/AP
Sample : VF1026SBL01
Misc : 5.00/5mL/MSVOA F/SOIL
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
VF1026SBL01

Quant Time: Oct 27 02:12:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F102518S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 26 02:17:41 2018
Response via : Initial Calibration

