

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_F\DATA\VF103018\
 Data File : VF060548.D
 Acq On : 30 Oct 2018 16:48
 Operator : VA/AP
 Sample : J5711-05
 Misc : 5.93/5mL/MSVOA_F/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SB-6

Manual Integrations
 APPROVED

apatel
 10/31/2018 10:29:22 AM

Quant Time: Oct 31 05:28:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F102918S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 30 05:31:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.04	168	173941	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	5.77	114	291199	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.91	117	209629	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	12.63	152	92338	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.02	65	194064	72.33	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	144.66%	
35) Dibromofluoromethane	4.29	113	171134	61.24	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	122.48%	
50) Toluene-d8	7.70	98	297868	43.65	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	87.30%	
62) 4-Bromofluorobenzene	11.51	95	121263	38.57	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	77.14%	
Target Compounds						
16) Acetone	2.34	43	26819	44.05	ug/l	# 85
20) Methylene Chloride	2.27	84	14725m	7.49	ug/l	
51) 4-Methyl-2-Pentanone	8.41	43	12964	12.59	ug/l	# 83
59) 2-Hexanone	9.62	43	22428	30.03	ug/l	95
64) Tetrachloroethene	8.30	164	13982	7.89	ug/l	87
95) Naphthalene	14.52	128	171172	44.23	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_F\DATA\VF103018\
Data File : VF060548.D
Acq On : 30 Oct 2018 16:48
Operator : VA/AP
Sample : J5711-05
Misc : 5.93/5mL/MSVOA_F/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleID :
SB-6

Manual Integrations
APPROVED

apatel
10/31/2018 10:29:22 AM

Quant Time: Oct 31 05:28:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F102918S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 30 05:31:58 2018
Response via : Initial Calibration

