

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_F\DATA\VF110218\  
 Data File : VF060561.D  
 Acq On : 2 Nov 2018 13:28  
 Operator : VA/AP  
 Sample : VSTDICC100  
 Misc : 5.00g/5mL/MSVOA-F/SOIL  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDICC100

Quant Time: Nov 03 04:31:01 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_F\METHODS\82F110218S.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Nov 03 04:25:32 2018  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 4.87  | 168  | 244374   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 5.60  | 114  | 432835   | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 9.78  | 117  | 378211   | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.56 | 152  | 203151   | 50.00 | ug/l  | 0.00     |

|                             |        |     |          |       |         |      |
|-----------------------------|--------|-----|----------|-------|---------|------|
| System Monitoring Compounds |        |     |          |       |         |      |
| 33) 1,2-Dichloroethane-d4   | 4.84   | 65  | 250178   | 70.23 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 140.46% |      |
| 35) Dibromofluoromethane    | 4.11   | 113 | 310723   | 76.46 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 152.92% |      |
| 50) Toluene-d8              | 7.55   | 98  | 890824   | 88.86 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 177.72% |      |
| 62) 4-Bromofluorobenzene    | 11.41  | 95  | 381996   | 83.34 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 166.68% |      |

| Target Compounds              |      |     |         |                | Qvalue |
|-------------------------------|------|-----|---------|----------------|--------|
| 2) Dichlorodifluoromethane    | 0.97 | 85  | 252781  | 55.556 ug/l    | 95     |
| 3) Chloromethane              | 1.10 | 50  | 191024  | 69.658 ug/l    | 96     |
| 4) Vinyl Chloride             | 1.14 | 62  | 180565  | 63.924 ug/l #  | 88     |
| 5) Bromomethane               | 1.33 | 94  | 105441  | 49.233 ug/l    | 98     |
| 6) Chloroethane               | 1.38 | 64  | 66512   | 46.209 ug/l    | 93     |
| 7) Trichlorofluoromethane     | 1.50 | 101 | 351048  | 53.329 ug/l    | 94     |
| 8) Diethyl Ether              | 1.64 | 74  | 60303   | 70.520 ug/l    | 95     |
| 9) 1,1,2-Trichlorotrifluoroet | 1.82 | 101 | 196768  | 60.053 ug/l    | 98     |
| 10) Methyl Iodide             | 1.89 | 142 | 269096  | 52.637 ug/l    | 98     |
| 11) Tert butyl alcohol        | 2.57 | 59  | 66105   | 386.912 ug/l # | 87     |
| 12) 1,1-Dichloroethene        | 1.78 | 96  | 168555  | 63.883 ug/l    | 95     |
| 13) Acrolein                  | 2.00 | 56  | 37784   | 257.816 ug/l   | 100    |
| 14) Allyl chloride            | 2.09 | 41  | 244981  | 57.826 ug/l    | 93     |
| 15) Acrylonitrile             | 2.95 | 53  | 138419  | 377.151 ug/l   | 98     |
| 16) Acetone                   | 2.24 | 43  | 254209  | 327.810 ug/l   | 97     |
| 17) Carbon Disulfide          | 1.83 | 76  | 356538  | 48.455 ug/l    | 99     |
| 18) Methyl Acetate            | 2.35 | 43  | 94469   | 83.732 ug/l    | 97     |
| 19) Methyl tert-butyl Ether   | 2.44 | 73  | 419866  | 72.518 ug/l    | 99     |
| 20) Methylene Chloride        | 2.24 | 84  | 107568  | 38.970 ug/l    | 89     |
| 21) trans-1,2-Dichloroethene  | 2.31 | 96  | 174868  | 64.728 ug/l    | 97     |
| 22) Diisopropyl ether         | 2.80 | 45  | 604399  | 78.741 ug/l    | 96     |
| 23) Vinyl Acetate             | 3.18 | 43  | 1775491 | 586.878 ug/l   | 99     |
| 24) 1,1-Dichloroethane        | 2.87 | 63  | 354074  | 64.460 ug/l    | 97     |
| 25) 2-Butanone                | 4.34 | 43  | 516439  | 456.847 ug/l   | 97     |
| 26) 2,2-Dichloropropane       | 3.60 | 77  | 245453  | 72.375 ug/l    | 99     |
| 27) cis-1,2-Dichloroethene    | 3.47 | 96  | 354545  | 103.954 ug/l   | 93     |
| 28) Bromochloromethane        | 3.72 | 49  | 215843  | 103.523 ug/l   | 95     |
| 29) Tetrahydrofuran           | 4.07 | 42  | 210737  | 601.191 ug/l   | 98     |
| 30) Chloroform                | 3.86 | 83  | 545871  | 78.004 ug/l    | 98     |
| 31) Cyclohexane               | 3.69 | 56  | 468799  | 110.724 ug/l # | 95     |
| 32) 1,1,1-Trichloroethane     | 4.10 | 97  | 326530  | 57.676 ug/l    | 98     |
| 36) 1,1-Dichloropropene       | 4.28 | 75  | 439345  | 82.806 ug/l    | 96     |
| 37) Ethyl Acetate             | 4.11 | 43  | 191428  | 105.652 ug/l   | 93     |
| 38) Carbon Tetrachloride      | 3.99 | 117 | 352658  | 58.175 ug/l    | 98     |

Data Path : Z:\VOASRV\HPCHEM1\MSVOA\_F\DATA\VF110218\  
 Data File : VF060561.D  
 Acq On : 2 Nov 2018 13:28  
 Operator : VA/AP  
 Sample : VSTDICC100  
 Misc : 5.00g/5mL/MSVOA-F/SOIL  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDICC100

Quant Time: Nov 03 04:31:01 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_F\METHODS\82F110218S.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Nov 03 04:25:32 2018  
 Response via : Initial Calibration

| Internal Standards             | R.T.  | QIon | Response | Conc     | Units  | Dev(Min) |
|--------------------------------|-------|------|----------|----------|--------|----------|
| 39) Methylcyclohexane          | 5.44  | 83   | 514212   | 92.372   | ug/l   | 99       |
| 40) Benzene                    | 4.64  | 78   | 1022263  | 94.266   | ug/l   | 98       |
| 41) Methacrylonitrile          | 4.74  | 41   | 109819   | 105.829  | ug/l # | 93       |
| 42) 1,2-Dichloroethane         | 4.94  | 62   | 318672   | 68.160   | ug/l   | 96       |
| 43) Isopropyl Acetate          | 6.96  | 43   | 262548   | 121.453  | ug/l # | 100      |
| 44) Trichloroethene            | 5.50  | 130  | 305901   | 87.760   | ug/l   | 98       |
| 45) 1,2-Dichloropropane        | 6.23  | 63   | 277154   | 102.304  | ug/l   | 100      |
| 46) Dibromomethane             | 6.09  | 93   | 183733   | 89.895   | ug/l   | 88       |
| 47) Bromodichloromethane       | 6.39  | 83   | 424757   | 82.023   | ug/l   | 97       |
| 48) Methyl methacrylate        | 6.72  | 41   | 163043   | 111.598  | ug/l   | 95       |
| 49) 1,4-Dioxane                | 6.71  | 88   | 32590    | 2286.310 | ug/l # | 85       |
| 51) 4-Methyl-2-Pentanone       | 8.27  | 43   | 801510   | 514.388  | ug/l   | 97       |
| 52) Toluene                    | 7.62  | 92   | 684671   | 98.381   | ug/l   | 95       |
| 53) t-1,3-Dichloropropene      | 8.28  | 75   | 340385   | 83.516   | ug/l   | 99       |
| 54) cis-1,3-Dichloropropene    | 7.30  | 75   | 453170   | 91.760   | ug/l   | 99       |
| 55) 1,1,2-Trichloroethane      | 8.49  | 97   | 203829   | 99.896   | ug/l   | 96       |
| 56) Ethyl methacrylate         | 8.62  | 69   | 246581   | 106.737  | ug/l   | 98       |
| 57) 1,3-Dichloropropane        | 8.86  | 76   | 349204   | 101.043  | ug/l   | 94       |
| 58) 2-Chloroethyl Vinyl ether  | 7.62  | 63   | 86089    | 370.838  | ug/l   | 100      |
| 59) 2-Hexanone                 | 9.50  | 43   | 604769   | 535.853  | ug/l   | 96       |
| 60) Dibromochloromethane       | 8.72  | 129  | 278940   | 85.935   | ug/l   | 99       |
| 61) 1,2-Dibromoethane          | 8.99  | 107  | 221518   | 94.191   | ug/l   | 100      |
| 64) Tetrachloroethene          | 8.16  | 164  | 280189   | 89.025   | ug/l   | 96       |
| 65) Chlorobenzene              | 9.80  | 112  | 727068   | 96.134   | ug/l   | 97       |
| 66) 1,1,1,2-Tetrachloroethane  | 9.93  | 131  | 298630   | 86.894   | ug/l   | 94       |
| 67) Ethyl Benzene              | 9.91  | 91   | 1142657  | 79.369   | ug/l   | 94       |
| 68) m/p-Xylenes                | 10.14 | 106  | 937047   | 188.397  | ug/l   | 89       |
| 69) o-Xylene                   | 10.71 | 106  | 521732   | 93.531   | ug/l   | 98       |
| 70) Styrene                    | 10.79 | 104  | 729030   | 92.339   | ug/l   | 96       |
| 71) Bromoform                  | 10.78 | 173  | 159816   | 91.321   | ug/l   | 99       |
| 73) Isopropylbenzene           | 11.13 | 105  | 1360325  | 86.572   | ug/l   | 100      |
| 74) N-amyl acetate             | 11.37 | 43   | 389137   | 118.218  | ug/l   | 97       |
| 75) 1,1,2,2-Tetrachloroethane  | 11.70 | 83   | 265131   | 100.823  | ug/l   | 98       |
| 76) 1,2,3-Trichloropropane     | 11.88 | 75   | 127054   | 70.377   | ug/l   | 98       |
| 77) Bromobenzene               | 11.50 | 156  | 314358   | 91.418   | ug/l   | 89       |
| 78) n-propylbenzene            | 11.60 | 91   | 1559442  | 84.219   | ug/l   | 92       |
| 79) 2-Chlorotoluene            | 11.73 | 91   | 910687   | 82.601   | ug/l   | 97       |
| 80) 1,3,5-Trimethylbenzene     | 11.83 | 105  | 1090839  | 81.382   | ug/l   | 99       |
| 81) trans-1,4-Dichloro-2-buten | 11.88 | 75   | 127054   | 144.144  | ug/l   | 98       |
| 82) 4-Chlorotoluene            | 11.90 | 91   | 954294   | 83.833   | ug/l   | 95       |
| 83) tert-Butylbenzene          | 12.13 | 119  | 1117414  | 84.841   | ug/l   | 96       |
| 84) 1,2,4-Trimethylbenzene     | 12.21 | 105  | 1120264  | 82.269   | ug/l   | 95       |
| 85) sec-Butylbenzene           | 12.31 | 105  | 1551997  | 84.066   | ug/l   | 94       |
| 86) p-Isopropyltoluene         | 12.47 | 119  | 1239950  | 81.902   | ug/l   | 96       |
| 87) 1,3-Dichlorobenzene        | 12.48 | 146  | 578533   | 84.175   | ug/l   | 98       |
| 88) 1,4-Dichlorobenzene        | 12.58 | 146  | 578749   | 89.104   | ug/l   | 97       |
| 89) n-Butylbenzene             | 12.85 | 91   | 1190941  | 75.034   | ug/l   | 95       |
| 90) Hexachloroethane           | 12.92 | 117  | 343840   | 88.122   | ug/l   | 87       |
| 91) 1,2-Dichlorobenzene        | 12.95 | 146  | 565239   | 87.683   | ug/l   | 97       |
| 92) 1,2-Dibromo-3-Chloropropan | 13.65 | 75   | 49037    | 92.101   | ug/l   | 89       |

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA\_F\DATA\VF110218\  
Data File : VF060561.D  
Acq On : 2 Nov 2018 13:28  
Operator : VA/AP  
Sample : VSTDICC100  
Misc : 5.00g/5mL/MSVOA-F/SOIL  
ALS Vial : 7 Sample Multiplier: 1

Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDICC100

Quant Time: Nov 03 04:31:01 2018  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_F\METHODS\82F110218S.M  
Quant Title : SW846 8260  
QLast Update : Sat Nov 03 04:25:32 2018  
Response via : Initial Calibration


| Internal Standards         | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|-------|------|----------|--------|-------|----------|
| 93) 1,2,4-Trichlorobenzene | 14.20 | 180  | 381070   | 79.886 | ug/l  | 95       |
| 94) Hexachlorobutadiene    | 14.19 | 225  | 252383   | 75.597 | ug/l  | 97       |
| 95) Naphthalene            | 14.46 | 128  | 830931   | 97.937 | ug/l  | 98       |
| 96) 1,2,3-Trichlorobenzene | 14.61 | 180  | 384566   | 84.929 | ug/l  | 94       |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA\_F\DATA\VF110218\  
 Data File : VF060561.D  
 Acq On : 2 Nov 2018 13:28  
 Operator : VA/AP  
 Sample : VSTDICC100  
 Misc : 5.00g/5mL/MSVOA-F/SOIL  
 ALS Vial : 7 Sample Multiplier: 1

Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDICC100

Quant Time: Nov 03 04:31:01 2018  
 Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_F\METHODS\82F110218S.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Nov 03 04:25:32 2018  
 Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.87 min Scan# 430

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :


VSTDIC100

Tgt Ion:168 Resp: 244374

Ion Ratio Lower Upper




168 100

99 59.1 48.6 72.8



Abundance Ion 168.00 (167.70 to 168.70): VF0

Ion 99.00 (98.70 to 99.70): VF0



#2

Dichlorodifluoromethane

Concen: 55.556 ug/l

RT: 0.97 min Scan# 34

Delta R.T. 0.00 min

Lab File: VF060561.D


Acq: 2 Nov 2018 13:28

Tgt Ion: 85 Resp: 252781

Ion Ratio Lower Upper



85 100


87 30.9 16.9 50.7



Abundance Ion 85.00 (84.70 to 85.70): VF0

Ion 87.00 (86.70 to 87.70): VF0





#3

Chloromethane

Concen: 69.658 ug/l

RT: 1.10 min Scan# 48

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :





VSTDIC100

Tgt Ion: 50 Resp: 191024

Ion Ratio Lower Upper

50 100

52 32.0 24.0 36.0



#4

Vinyl Chloride

Concen: 63.924 ug/l

RT: 1.14 min Scan# 52

Delta R.T. 0.00 min

Lab File: VF060561.D




Acq: 2 Nov 2018 13:28


Tgt Ion: 62 Resp: 180565

Ion Ratio Lower Upper

62 100

64 33.4 21.8 32.6#





#5

Bromomethane

Concen: 49.233 ug/l

RT: 1.33 min Scan# 71

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :





VSTDIC100

Tgt Ion: 94 Resp: 105441

Ion Ratio Lower Upper

94 100

96 95.3 77.9 116.9



#6

Chloroethane

Concen: 46.209 ug/l

RT: 1.38 min Scan# 76

Delta R.T. 0.00 min

Lab File: VF060561.D




Acq: 2 Nov 2018 13:28

Tgt Ion: 64 Resp: 66512


Ion Ratio Lower Upper

64 100

66 30.6 27.8 41.8







#7

Trichlorofluoromethane

Concen: 53.329 ug/l

RT: 1.50 min Scan# 88

Delta R.T. 0.01 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :


VSTDIC100

Tgt Ion:101 Resp: 351048

Ion Ratio Lower Upper




101 100

103 64.5 48.0 72.0



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 103.00 (102.70 to 103.70): V



#8

Diethyl Ether

Concen: 70.520 ug/l

RT: 1.64 min Scan# 102

Delta R.T. 0.00 min

Lab File: VF060561.D


Acq: 2 Nov 2018 13:28

Tgt Ion: 74 Resp: 60303

Ion Ratio Lower Upper



74 100

45 133.2 63.8 191.5




Abundance Ion 74.00 (73.70 to 74.70): VF0

Ion 45.00 (44.70 to 45.70): VF0







#9

1,1,2-Trichlorotrifluoroethane

Concen: 60.053 ug/l

RT: 1.82 min Scan# 121

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC100


Tgt Ion:101 Resp: 196768

Ion Ratio Lower Upper

101 100

85 49.1 39.8 59.6




151 65.7 50.8 76.2



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 85.00 (84.70 to 85.70): VF0

Ion 151.00 (150.70 to 151.70): V



#10

Methyl Iodide

Concen: 52.637 ug/l

RT: 1.89 min Scan# 128

Delta R.T. -0.01 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28


Tgt Ion:142 Resp: 269096

Ion Ratio Lower Upper

142 100

127 61.2 47.8 71.6



141 15.3 12.7 19.1




Abundance Ion 142.00 (141.70 to 142.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 141.00 (140.70 to 141.70): V









#11  
Tert butyl alcohol  
Concen: 386.912 ug/l  
RT: 2.57 min Scan# 197  
Delta R.T. 0.00 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28




Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC100


Tgt Ion: 59 Resp: 66105  
Ion Ratio Lower Upper  
59 100  
57 17.5 19.0 28.4#



#12  
1,1-Dichloroethene  
Concen: 63.883 ug/l  
RT: 1.78 min Scan# 117  
Delta R.T. 0.00 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28

Tgt Ion: 96 Resp: 168555  
Ion Ratio Lower Upper  
96 100  
61 190.8 158.1 237.1  
98 63.7 47.0 70.6





#13

Acrolein

Concen: 257.816 ug/l

RT: 2.00 min Scan# 139

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :





VSTDIC100

Tgt Ion: 56 Resp: 37784

Ion Ratio Lower Upper

56 100

55 76.9 61.8 92.6



#14

Allyl chloride

Concen: 57.826 ug/l

RT: 2.09 min Scan# 148

Delta R.T. -0.01 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28




Tgt Ion: 41 Resp: 244981


Ion Ratio Lower Upper

41 100

39 96.6 82.6 124.0

76 29.3 26.7 40.1









#15  
 Acrylonitrile  
 Concen: 377.151 ug/l  
 RT: 2.95 min Scan# 235  
 Delta R.T. 0.00 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28




Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC100


Tgt Ion: 53 Resp: 138419  
 Ion Ratio Lower Upper  
 53 100  
 52 86.6 68.5 102.7  
 51 50.4 38.2 57.2



#16  
 Acetone  
 Concen: 327.810 ug/l  
 RT: 2.24 min Scan# 163  
 Delta R.T. 0.00 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28

Tgt Ion: 43 Resp: 254209  
 Ion Ratio Lower Upper  
 43 100  
 58 16.9 14.6 22.0





#17

Carbon Disulfide

Concen: 48.455 ug/l

RT: 1.83 min Scan# 122

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :





VSTDIC100

Tgt Ion: 76 Resp: 356538

Ion Ratio Lower Upper

76 100

78 12.0 9.9 14.9



#18

Methyl Acetate

Concen: 83.732 ug/l

RT: 2.35 min Scan# 174

Delta R.T. 0.00 min

Lab File: VF060561.D




Acq: 2 Nov 2018 13:28


Tgt Ion: 43 Resp: 94469

Ion Ratio Lower Upper

43 100

74 16.5 12.1 18.1









#19

Methyl tert-butyl Ether  
 Concen: 72.518 ug/l  
 RT: 2.44 min Scan# 183  
 Delta R.T. 0.00 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28

Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC100




Tgt Ion: 73 Resp: 419866  
 Ion Ratio Lower Upper  
 73 100  
 57 19.5 15.4 23.0




#20

Methylene Chloride  
 Concen: 38.970 ug/l  
 RT: 2.24 min Scan# 163  
 Delta R.T. 0.00 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28

Tgt Ion: 84 Resp: 107568  
 Ion Ratio Lower Upper  
 84 100  
 49 154.3 109.9 164.9  
 51 47.8 35.8 53.8  
 86 65.9 48.3 72.5





#21

trans-1,2-Dichloroethene

Concen: 64.728 ug/l

RT: 2.31 min Scan# 170

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC100





Tgt Ion: 96 Resp: 174868

Ion Ratio Lower Upper

96 100

61 159.0 131.3 196.9

98 62.0 48.6 73.0



#22

Diisopropyl ether

Concen: 78.741 ug/l

RT: 2.80 min Scan# 220

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Tgt Ion: 45 Resp: 604399




Ion Ratio Lower Upper

45 100


43 44.8 38.1 57.1

87 17.7 12.7 19.1

59 9.3 7.0 10.4







#23

Vinyl Acetate

Concen: 586.878 ug/l

RT: 3.18 min Scan# 259

Delta R.T. -0.01 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :


VSTDIC100

Tgt Ion: 43 Resp: 1775491

Ion Ratio Lower Upper




43 100

86 9.3 7.1 10.7



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 86.00 (85.70 to 86.70): VF0



#24

1,1-Dichloroethane

Concen: 64.460 ug/l

RT: 2.87 min Scan# 227

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28


Tgt Ion: 63 Resp: 354074

Ion Ratio Lower Upper

63 100

98 5.7 2.2 6.6



100 3.1 1.4 4.0




Abundance Ion 63.00 (62.70 to 63.70): VF0

Ion 98.00 (97.70 to 98.70): VF0

Ion 100.00 (99.70 to 100.70): VF0





#25

2-Butanone

Concen: 456.847 ug/l

RT: 4.34 min Scan# 376

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :





VSTDIC100

Tgt Ion: 43 Resp: 516439

Ion Ratio Lower Upper

43 100

72 18.7 16.2 24.2



#26

2,2-Dichloropropane

Concen: 72.375 ug/l

RT: 3.60 min Scan# 301

Delta R.T. 0.00 min

Lab File: VF060561.D




Acq: 2 Nov 2018 13:28


Tgt Ion: 77 Resp: 245453

Ion Ratio Lower Upper

77 100

97 19.9 9.8 29.3









#27

cis-1,2-Dichloroethene  
 Concen: 103.954 ug/l  
 RT: 3.47 min Scan# 288  
 Delta R.T. -0.01 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28

Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC100




| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 96      | 100   |       |       |
| 61      | 119.9 | 0.0   | 263.0 |
| 98      | 65.8  | 0.0   | 130.6 |




#28

Bromochloromethane  
 Concen: 103.523 ug/l  
 RT: 3.72 min Scan# 313  
 Delta R.T. 0.00 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 49      | 100   |       |       |
| 129     | 0.0   | 0.0   | 4.6   |
| 130     | 81.2  | 61.8  | 92.8  |





#29

Tetrahydrofuran

Concen: 601.191 ug/l

RT: 4.07 min Scan# 349

Delta R.T. -0.01 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC100


Tgt Ion: 42 Resp: 210737

Ion Ratio Lower Upper

42 100

72 38.9 31.4 47.0




71 37.9 31.8 47.6



Abundance Ion 42.00 (41.70 to 42.70): VF0

Ion 72.00 (71.70 to 72.70): VF0

Ion 71.00 (70.70 to 71.70): VF0



#30

Chloroform

Concen: 78.004 ug/l

RT: 3.86 min Scan# 327

Delta R.T. 0.00 min

Lab File: VF060561.D


Acq: 2 Nov 2018 13:28

Tgt Ion: 83 Resp: 545871

Ion Ratio Lower Upper



83 100


85 65.8 53.7 80.5



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 85.00 (84.70 to 85.70): VF0









#31  
Cyclohexane  
Concen: 110.724 ug/l  
RT: 3.69 min Scan# 310  
Delta R.T. 0.00 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28




Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC100


Tgt Ion: 56 Resp: 468799  
Ion Ratio Lower Upper  
56 100  
69 34.8 23.0 34.4#  
84 87.0 67.1 100.7



#32  
1,1,1-Trichloroethane  
Concen: 57.676 ug/l  
RT: 4.10 min Scan# 352  
Delta R.T. 0.01 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28

Tgt Ion: 97 Resp: 326530  
Ion Ratio Lower Upper  
97 100  
99 64.3 52.4 78.6  
61 49.6 41.5 62.3









#33  
 1,2-Dichloroethane-d4  
 Concen: 70.232 ug/l  
 RT: 4.84 min Scan# 427  
 Delta R.T. -0.01 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28




Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC100


Tgt Ion: 65 Resp: 250178  
 Ion Ratio Lower Upper  
 65 100  
 67 51.4 0.0 101.8



#34  
 1,4-Difluorobenzene  
 Concen: 50.000 ug/l  
 RT: 5.60 min Scan# 504  
 Delta R.T. 0.00 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28

Tgt Ion: 114 Resp: 432835  
 Ion Ratio Lower Upper  
 114 100  
 63 18.6 0.0 39.0  
 88 18.4 0.0 38.2









#35  
 Dibromofluoromethane  
 Concen: 76.457 ug/l  
 RT: 4.11 min Scan# 353  
 Delta R.T. 0.00 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28

Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC100

Tgt Ion: 113 Resp: 310723  
 Ion Ratio Lower Upper  
 113 100  
 111 101.4 79.9 119.9  
 192 17.5 13.3 19.9




Abundance Ion 113.00 (112.70 to 113.70): V  
 Ion 111.00 (110.70 to 111.70): V  
 Ion 192.00 (191.70 to 192.70): V






#36  
 1,1-Dichloropropene  
 Concen: 82.806 ug/l  
 RT: 4.28 min Scan# 370  
 Delta R.T. 0.00 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28

Tgt Ion: 75 Resp: 439345  
 Ion Ratio Lower Upper  
 75 100  
 110 35.8 16.2 48.6  
 77 32.6 25.0 37.6



Abundance Ion 75.00 (74.70 to 75.70): VF0  
 Ion 110.00 (109.70 to 110.70): V  
 Ion 77.00 (76.70 to 77.70): VF0





#37

Ethyl Acetate

Concen: 105.652 ug/l

RT: 4.11 min Scan# 353

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC100


Tgt Ion: 43 Resp: 191428

Ion Ratio Lower Upper




43 100

61 72.4 63.4 95.2

70 10.5 9.3 13.9



Abundance Ion 43.00 (42.70 to 43.70): VF0  
Ion 61.00 (60.70 to 61.70): VF0  
Ion 70.00 (69.70 to 70.70): VF0



#38

Carbon Tetrachloride

Concen: 58.175 ug/l

RT: 3.99 min Scan# 341

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28


Tgt Ion: 117 Resp: 352658

Ion Ratio Lower Upper



117 100

119 98.4 77.8 116.6


121 30.0 25.6 38.4



Abundance Ion 117.00 (116.70 to 117.70): V  
Ion 119.00 (118.70 to 119.70): V  
Ion 121.00 (120.70 to 121.70): V











#39  
Methylcyclohexane  
Concen: 92.372 ug/l  
RT: 5.44 min Scan# 488  
Delta R.T. -0.01 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28




Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC100


Tgt Ion: 83 Resp: 514212  
Ion Ratio Lower Upper  
83 100  
55 85.1 69.1 103.7  
98 49.0 38.6 57.8



#40  
Benzene  
Concen: 94.266 ug/l  
RT: 4.64 min Scan# 406  
Delta R.T. 0.00 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28

Tgt Ion: 78 Resp: 1022263  
Ion Ratio Lower Upper  
78 100  
77 22.9 19.3 28.9









#41  
Methacrylonitrile  
Concen: 105.829 ug/l  
RT: 4.74 min Scan# 417  
Delta R.T. -0.01 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28




Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC100


Tgt Ion: 41 Resp: 109819  
Ion Ratio Lower Upper  
41 100  
39 53.3 45.6 68.4  
67 53.7 45.3 67.9  
52 45.5 29.6 44.4#



#42  
1,2-Dichloroethane  
Concen: 68.160 ug/l  
RT: 4.94 min Scan# 437  
Delta R.T. -0.01 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28

Tgt Ion: 62 Resp: 318672  
Ion Ratio Lower Upper  
62 100  
98 9.9 0.0 22.4





#43

Isopropyl Acetate

Concen: 121.453 ug/l

RT: 6.96 min Scan# 642

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC100


Tgt Ion: 43 Resp: 262548

Ion Ratio Lower Upper

43 100

61 30.4 24.2 36.2




87 0.5 0.2 0.4#



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 87.00 (86.70 to 87.70): VF0



#44

Trichloroethene

Concen: 87.760 ug/l

RT: 5.50 min Scan# 494

Delta R.T. 0.00 min

Lab File: VF060561.D


Acq: 2 Nov 2018 13:28

Tgt Ion: 130 Resp: 305901

Ion Ratio Lower Upper



130 100


95 97.7 0.0 191.0



Abundance Ion 129.90 (129.60 to 130.60): V

Ion 94.90 (94.60 to 95.60): VF0









#45  
 1,2-Dichloropropane  
 Concen: 102.304 ug/l  
 RT: 6.23 min Scan# 568  
 Delta R.T. -0.01 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28




Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC100


Tgt Ion: 63 Resp: 277154  
 Ion Ratio Lower Upper  
 63 100  
 65 31.5 25.4 38.0



#46  
 Dibromomethane  
 Concen: 89.895 ug/l  
 RT: 6.09 min Scan# 553  
 Delta R.T. -0.01 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28

Tgt Ion: 93 Resp: 183733  
 Ion Ratio Lower Upper  
 93 100  
 95 78.4 68.6 102.8  
 174 83.8 79.0 118.4





#47

Bromodichloromethane

Concen: 82.023 ug/l

RT: 6.39 min Scan# 584

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC100


Tgt Ion: 83 Resp: 424757

Ion Ratio Lower Upper

83 100

85 67.3 51.8 77.6




127 8.6 7.0 10.4



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 85.00 (84.70 to 85.70): VF0

Ion 127.00 (126.70 to 127.70): V



#48

Methyl methacrylate

Concen: 111.598 ug/l

RT: 6.72 min Scan# 617

Delta R.T. -0.01 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28


Tgt Ion: 41 Resp: 163043

Ion Ratio Lower Upper

41 100

69 72.2 61.1 91.7



39 57.4 43.0 64.6




Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 69.00 (68.70 to 69.70): VF0

Ion 39.00 (38.70 to 39.70): VF0









#49  
1,4-Dioxane  
Concen: 2286.310 ug/l  
RT: 6.71 min Scan# 616  
Delta R.T. -0.01 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28

Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC100

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 88      | 100   |       |       |
| 43      | 55.4  | 61.2  | 91.8# |
| 58      | 68.0  | 56.9  | 85.3  |




Abundance Ion 88.00 (87.70 to 88.70): VF0  
Ion 43.00 (42.70 to 43.70): VF0  
Ion 58.00 (57.70 to 58.70): VF0






#50  
Toluene-d8  
Concen: 88.856 ug/l  
RT: 7.55 min Scan# 702  
Delta R.T. -0.01 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 98      | 100   |       |       |
| 100     | 76.4  | 59.4  | 89.2  |



Abundance Ion 98.00 (97.70 to 98.70): VF0  
Ion 100.00 (99.70 to 100.70): VF0









#51  
 4-Methyl-2-Pentanone  
 Concen: 514.388 ug/l  
 RT: 8.27 min Scan# 774  
 Delta R.T. 0.00 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28




Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC100


Tgt Ion: 43 Resp: 801510  
 Ion Ratio Lower Upper  
 43 100  
 58 39.4 30.1 45.1



#52  
 Toluene  
 Concen: 98.381 ug/l  
 RT: 7.62 min Scan# 709  
 Delta R.T. -0.01 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28

Tgt Ion: 92 Resp: 684671  
 Ion Ratio Lower Upper  
 92 100  
 91 149.5 124.5 186.7





#53

t-1,3-Dichloropropene

Concen: 83.516 ug/l

RT: 8.28 min Scan# 776

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :





VSTDIC100

Tgt Ion: 75 Resp: 340385

Ion Ratio Lower Upper

75 100

77 32.9 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 91.760 ug/l

RT: 7.30 min Scan# 676

Delta R.T. -0.01 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28




Tgt Ion: 75 Resp: 453170

Ion Ratio Lower Upper


75 100

77 33.4 26.7 40.1

39 55.3 43.0 64.4







#55

1,1,2-Trichloroethane

Concen: 99.896 ug/l

RT: 8.49 min Scan# 797

Delta R.T. -0.01 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC100

Tgt Ion: 97 Resp: 203829


Ion Ratio Lower Upper

97 100

83 83.2 69.4 104.0

85 55.4 46.6 69.8

99 60.0 50.0 75.0






Abundance Ion 97.00 (96.70 to 97.70): VF0

Ion 82.95 (82.65 to 83.65): VF0

Ion 84.95 (84.65 to 85.65): VF0

Ion 98.95 (98.65 to 99.65): VF0



#56

Ethyl methacrylate

Concen: 106.737 ug/l

RT: 8.62 min Scan# 810

Delta R.T. -0.01 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28


Tgt Ion: 69 Resp: 246581

Ion Ratio Lower Upper

69 100

41 75.1 59.2 88.8



39 38.1 32.3 48.5




Abundance Ion 69.00 (68.70 to 69.70): VF0

Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0









#57  
 1,3-Dichloropropane  
 Concen: 101.043 ug/l  
 RT: 8.86 min Scan# 834  
 Delta R.T. 0.00 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28




Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC100


Tgt Ion: 76 Resp: 349204  
 Ion Ratio Lower Upper  
 76 100  
 78 33.6 24.2 36.4



#58  
 2-Chloroethyl Vinyl ether  
 Concen: 370.838 ug/l  
 RT: 7.62 min Scan# 709  
 Delta R.T. -0.01 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28

Tgt Ion: 63 Resp: 86089  
 Ion Ratio Lower Upper  
 63 100  
 106 0.0 0.0 0.0









#59  
2-Hexanone  
Concen: 535.853 ug/l  
RT: 9.50 min Scan# 899  
Delta R.T. 0.00 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28




Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC100


Tgt Ion: 43 Resp: 604769  
Ion Ratio Lower Upper  
43 100  
58 55.3 26.1 78.3



#60  
Dibromochloromethane  
Concen: 85.935 ug/l  
RT: 8.72 min Scan# 820  
Delta R.T. 0.00 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28

Tgt Ion: 129 Resp: 278940  
Ion Ratio Lower Upper  
129 100  
127 76.2 37.8 113.3





#61

1,2-Dibromoethane

Concen: 94.191 ug/l

RT: 8.99 min Scan# 847

Delta R.T. -0.01 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :


VSTDIC100

Tgt Ion: 107 Resp: 221518

Ion Ratio Lower Upper




107 100

109 97.2 77.5 116.3



Abundance Ion 107.00 (106.70 to 107.70): V

Ion 109.00 (108.70 to 109.70): V



#62

4-Bromofluorobenzene

Concen: 83.344 ug/l

RT: 11.41 min Scan# 1093

Delta R.T. -0.01 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28


Tgt Ion: 95 Resp: 381996

Ion Ratio Lower Upper

95 100

174 68.0 0.0 149.0



176 64.9 0.0 145.6




Abundance Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 9.78 min Scan# 928  
Delta R.T. 0.00 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28




Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC100


Tgt Ion:117 Resp: 378211  
Ion Ratio Lower Upper  
117 100  
82 54.8 43.9 65.9  
119 33.7 27.4 41.0



#64  
Tetrachloroethene  
Concen: 89.025 ug/l  
RT: 8.16 min Scan# 763  
Delta R.T. 0.00 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28

Tgt Ion:164 Resp: 280189  
Ion Ratio Lower Upper  
164 100  
166 133.0 101.5 152.3  
129 94.3 72.5 108.7  
131 92.4 76.4 114.6









#65  
Chlorobenzene  
Concen: 96.134 ug/l  
RT: 9.80 min Scan# 930  
Delta R.T. 0.00 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28




Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDICC100


Tgt Ion:112 Resp: 727068  
Ion Ratio Lower Upper  
112 100  
114 34.7 26.6 40.0



#66  
1,1,1,2-Tetrachloroethane  
Concen: 86.894 ug/l  
RT: 9.93 min Scan# 943  
Delta R.T. -0.01 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28

Tgt Ion:131 Resp: 298630  
Ion Ratio Lower Upper  
131 100  
133 93.7 49.5 148.7  
119 65.7 35.2 105.6





#67

Ethyl Benzene

Concen: 79.369 ug/l

RT: 9.91 min Scan# 941

Delta R.T. 0.01 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :





VSTDIC100

Tgt Ion: 91 Resp: 1142657

Ion Ratio Lower Upper

91 100

106 33.8 24.2 36.4



#68

m/p-Xylenes

Concen: 188.397 ug/l

RT: 10.14 min Scan# 964

Delta R.T. 0.00 min

Lab File: VF060561.D




Acq: 2 Nov 2018 13:28


Tgt Ion: 106 Resp: 937047

Ion Ratio Lower Upper

106 100

91 175.7 153.9 230.9









#69  
o-Xylene  
Concen: 93.531 ug/l  
RT: 10.71 min Scan# 1022  
Delta R.T. -0.01 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28




Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDICC100

Tgt Ion:106 Resp: 521732  
Ion Ratio Lower Upper  
106 100  
91 199.7 101.8 305.3




#70  
Styrene  
Concen: 92.339 ug/l  
RT: 10.79 min Scan# 1030  
Delta R.T. -0.01 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28

Tgt Ion:104 Resp: 729030  
Ion Ratio Lower Upper  
104 100  
78 44.8 32.6 49.0  
103 43.3 35.2 52.8











#71  
Bromoform  
Concen: 91.321 ug/l  
RT: 10.78 min Scan# 1029  
Delta R.T. 0.00 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28




Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC100


Tgt Ion:173 Resp: 159816  
Ion Ratio Lower Upper  
173 100  
175 48.4 24.3 72.9  
254 9.6 6.7 10.1



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 12.56 min Scan# 1209  
Delta R.T. 0.00 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28

Tgt Ion:152 Resp: 203151  
Ion Ratio Lower Upper  
152 100  
115 54.5 29.6 88.8  
150 154.4 0.0 315.6





#73

Isopropylbenzene

Concen: 86.572 ug/l

RT: 11.13 min Scan# 1064

Delta R.T. -0.01 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :





VSTDIC100

Tgt Ion:105 Resp: 1360325

Ion Ratio Lower Upper

105 100

120 27.6 13.9 41.6



#74

N-ethyl acetate

Concen: 118.218 ug/l

RT: 11.37 min Scan# 1089

Delta R.T. -0.01 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Tgt Ion: 43 Resp: 389137




Ion Ratio Lower Upper


43 100

70 35.2 30.9 46.3

55 20.3 16.6 24.8

61 26.6 22.0 33.0





#75

1,1,2,2-Tetrachloroethane

Concen: 100.823 ug/l

RT: 11.70 min Scan# 1122

Delta R.T. -0.01 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC100


Tgt Ion: 83 Resp: 265131

Ion Ratio Lower Upper

83 100

131 9.2 5.0 14.9




85 64.3 32.9 98.7



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 131.00 (130.70 to 131.70): VF0

Ion 85.00 (84.70 to 85.70): VF0



#76

1,2,3-Trichloropropane

Concen: 70.377 ug/l

RT: 11.88 min Scan# 1140

Delta R.T. 0.00 min

Lab File: VF060561.D


Acq: 2 Nov 2018 13:28

Tgt Ion: 75 Resp: 127054

Ion Ratio Lower Upper



75 100


77 36.0 17.4 52.3



Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 77.00 (76.70 to 77.70): VF0





#77

Bromobenzene

Concen: 91.418 ug/l

RT: 11.50 min Scan# 1102

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC100


Tgt Ion:156 Resp: 314358

Ion Ratio Lower Upper

156 100

77 155.5 89.3 268.1




158 101.2 50.0 150.1



Abundance Ion 156.00 (155.70 to 156.70): V

Ion 77.00 (76.70 to 77.70): VF0

Ion 158.00 (157.70 to 158.70): V



#78

n-propylbenzene

Concen: 84.219 ug/l

RT: 11.60 min Scan# 1112

Delta R.T. 0.00 min

Lab File: VF060561.D


Acq: 2 Nov 2018 13:28

Tgt Ion: 91 Resp: 1559442

Ion Ratio Lower Upper



91 100


120 24.7 10.5 31.6



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 120.00 (119.70 to 120.70): V









#79  
 2-Chlorotoluene  
 Concen: 82.601 ug/l  
 RT: 11.73 min Scan# 1125  
 Delta R.T. 0.00 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28




Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC100


Tgt Ion: 91 Resp: 910687  
 Ion Ratio Lower Upper  
 91 100  
 126 34.8 16.5 49.5



#80  
 1,3,5-Trimethylbenzene  
 Concen: 81.382 ug/l  
 RT: 11.83 min Scan# 1135  
 Delta R.T. -0.01 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28

Tgt Ion: 105 Resp: 1090839  
 Ion Ratio Lower Upper  
 105 100  
 120 55.8 27.6 82.8





#81

trans-1,4-Dichloro-2-butene

Concen: 144.144 ug/l

RT: 11.88 min Scan# 1140

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC100


Tgt Ion: 75 Resp: 127054

Ion Ratio Lower Upper

75 100

53 63.9 51.6 77.4




89 48.3 36.7 55.1



Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 53.00 (52.70 to 53.70): VF0

Ion 89.00 (88.70 to 89.70): VF0



#82

4-Chlorotoluene

Concen: 83.833 ug/l

RT: 11.90 min Scan# 1143

Delta R.T. 0.00 min

Lab File: VF060561.D


Acq: 2 Nov 2018 13:28

Tgt Ion: 91 Resp: 954294

Ion Ratio Lower Upper



91 100


126 34.7 15.9 47.7



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 126.00 (125.70 to 126.70): V









#83  
tert-Butylbenzene  
Concen: 84.841 ug/l  
RT: 12.13 min Scan# 1166  
Delta R.T. -0.01 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28

Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC100

Tgt Ion:119 Resp: 1117414  
Ion Ratio Lower Upper  
119 100  
91 61.5 28.8 86.4  
134 27.2 13.2 39.6




Abundance Ion 119.00 (118.70 to 119.70): V  
Ion 91.00 (90.70 to 91.70): VF0  
Ion 134.00 (133.70 to 134.70): V






#84  
1,2,4-Trimethylbenzene  
Concen: 82.269 ug/l  
RT: 12.21 min Scan# 1174  
Delta R.T. 0.00 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28

Tgt Ion:105 Resp: 1120264  
Ion Ratio Lower Upper  
105 100  
120 54.8 25.6 76.8



Abundance Ion 105.00 (104.70 to 105.70): V  
Ion 120.00 (119.70 to 120.70): V





#85

sec-Butylbenzene

Concen: 84.066 ug/l

RT: 12.31 min Scan# 1184

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :


VSTDICC100

Tgt Ion:105 Resp: 1551997

Ion Ratio Lower Upper




105 100

134 22.0 9.7 29.1



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 81.902 ug/l

RT: 12.47 min Scan# 1200

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28


Tgt Ion:119 Resp: 1239950

Ion Ratio Lower Upper

119 100

134 29.5 13.6 40.7



91 24.0 11.4 34.2




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): V







#87

1,3-Dichlorobenzene

Concen: 84.175 ug/l

RT: 12.48 min Scan# 1201

Delta R.T. -0.01 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC100


Tgt Ion:146 Resp: 578533

Ion Ratio Lower Upper

146 100

111 40.2 18.3 54.8




148 67.2 33.5 100.4



Abundance Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V



#88

1,4-Dichlorobenzene

Concen: 89.104 ug/l

RT: 12.58 min Scan# 1211

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28


Tgt Ion:146 Resp: 578749

Ion Ratio Lower Upper

146 100

111 36.4 19.7 59.1



148 69.0 33.7 101.1




Abundance Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V





#89

n-Butylbenzene

Concen: 75.034 ug/l

RT: 12.85 min Scan# 1239

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC100





Tgt Ion: 91 Resp: 1190941

Ion Ratio Lower Upper

91 100

92 71.2 34.1 102.2

134 32.7 13.9 41.6



#90

Hexachloroethane

Concen: 88.122 ug/l

RT: 12.92 min Scan# 1246

Delta R.T. -0.01 min

Lab File: VF060561.D




Acq: 2 Nov 2018 13:28


Tgt Ion: 117 Resp: 343840

Ion Ratio Lower Upper

117 100

201 59.4 35.2 105.6





#91

1,2-Dichlorobenzene

Concen: 87.683 ug/l

RT: 12.95 min Scan# 1249

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28

Instrument :

MSVOA\_F

ClientSampleId :

VSTDIC100


Tgt Ion:146 Resp: 565239

Ion Ratio Lower Upper

146 100

111 41.3 21.3 64.0

148 68.8 32.9 98.7






Abundance

Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 92.101 ug/l

RT: 13.65 min Scan# 1320

Delta R.T. 0.00 min

Lab File: VF060561.D

Acq: 2 Nov 2018 13:28


Tgt Ion: 75 Resp: 49037

Ion Ratio Lower Upper

75 100

155 98.5 47.3 141.8

157 128.7 55.3 165.9






Abundance

Ion 75.00 (74.70 to 75.70): VF0

Ion 155.00 (154.70 to 155.70): V

Ion 157.00 (156.70 to 157.70): V









#93  
 1,2,4-Trichlorobenzene  
 Concen: 79.886 ug/l  
 RT: 14.20 min Scan# 1376  
 Delta R.T. -0.01 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28

Instrument :  
 MSVOA\_F  
 ClientSampleId :  
 VSTDIC100

Tgt Ion:180 Resp: 381070  
 Ion Ratio Lower Upper  
 180 100  
 182 97.4 49.7 149.1  
 145 40.5 16.9 50.7




Abundance Ion 180.00 (179.70 to 180.70): V  
 Ion 182.00 (181.70 to 182.70): V  
 Ion 145.00 (144.70 to 145.70): V






#94  
 Hexachlorobutadiene  
 Concen: 75.597 ug/l  
 RT: 14.19 min Scan# 1375  
 Delta R.T. -0.01 min  
 Lab File: VF060561.D  
 Acq: 2 Nov 2018 13:28

Tgt Ion:225 Resp: 252383  
 Ion Ratio Lower Upper  
 225 100  
 223 63.0 30.9 92.5  
 227 63.9 33.9 101.6



Abundance Ion 225.00 (224.70 to 225.70): V  
 Ion 223.00 (222.70 to 223.70): V  
 Ion 227.00 (226.70 to 227.70): V









#95  
Naphthalene  
Concen: 97.937 ug/l  
RT: 14.46 min Scan# 1402  
Delta R.T. -0.01 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28

Instrument :  
MSVOA\_F  
ClientSampleId :  
VSTDIC100

Tgt Ion:128 Resp: 830931  
Ion Ratio Lower Upper  
128 100  
127 12.1 9.4 14.0  
129 11.9 8.4 12.6



#96  
1,2,3-Trichlorobenzene  
Concen: 84.929 ug/l  
RT: 14.61 min Scan# 1417  
Delta R.T. -0.01 min  
Lab File: VF060561.D  
Acq: 2 Nov 2018 13:28

Tgt Ion:180 Resp: 384566  
Ion Ratio Lower Upper  
180 100  
182 99.5 47.3 142.0  
145 35.5 15.6 46.8

