

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF110518\
 Data File : VF060581.D
 Acq On : 5 Nov 2018 19:43
 Operator : VA/AP
 Sample : J5741-19
 Misc : 6.72/5mL/MSVOA-F/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 TP-5

Quant Time: Nov 06 07:07:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 03 05:29:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.87	168	249271	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.60	114	482739	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.77	117	377127	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.55	152	161546	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.84	65	118631	47.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.20%	
35) Dibromofluoromethane	4.11	113	162390	43.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.08%	
50) Toluene-d8	7.55	98	414022	37.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.94%	
62) 4-Bromofluorobenzene	11.41	95	162576	33.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.28%	
Target Compounds						
20) Methylene Chloride	2.18	84	18205	5.480	ug/l	95
52) Toluene	7.62	92	27318	3.250	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF110518\
Data File : VF060581.D
Acq On : 5 Nov 2018 19:43
Operator : VA/AP
Sample : J5741-19
Misc : 6.72/5mL/MSVOA-F/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
TP-5

Quant Time: Nov 06 07:07:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 03 05:29:31 2018
Response via : Initial Calibration

