

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF110518\
 Data File : VF060582.D
 Acq On : 5 Nov 2018 20:12
 Operator : VA/AP
 Sample : J5741-23RE
 Misc : 6.58/5mL/MSVOA-F/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 TP-6RE

Quant Time: Nov 06 07:09:37 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 03 05:29:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.87	168	243734	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.60	114	460226	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.77	117	388263	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.55	152	203600	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.84	65	106684	43.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.64%	
35) Dibromofluoromethane	4.11	113	143873	40.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.92%	
50) Toluene-d8	7.55	98	321053	30.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	60.96%	
62) 4-Bromofluorobenzene	11.41	95	157571	34.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.40%	
Target Compounds						
16) Acetone	2.24	43	50852	87.632	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF110518\
Data File : VF060582.D
Acq On : 5 Nov 2018 20:12
Operator : VA/AP
Sample : J5741-23RE
Misc : 6.58/5mL/MSVOA-F/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
TP-6RE

Quant Time: Nov 06 07:09:37 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 03 05:29:31 2018
Response via : Initial Calibration

