

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF110518\
 Data File : VF060586.D
 Acq On : 5 Nov 2018 22:06
 Operator : VA/AP
 Sample : J5741-43
 Misc : 6.98/5mL/MSVOA-F/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 TP-11

Manual Integrations
 APPROVED

MMDadoda
 11/6/2018 12:53:49 PM

Quant Time: Nov 06 07:28:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 03 05:29:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.87	168	287968	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.60	114	522954	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.78	117	460843	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.56	152	213548	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.84	65	132612	45.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.16%	
35) Dibromofluoromethane	4.11	113	170356	42.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.32%	
50) Toluene-d8	7.56	98	444577	37.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.30%	
62) 4-Bromofluorobenzene	11.41	95	200286	38.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.52%	
Target Compounds						
16) Acetone	2.24	43	28441	41.483	ug/l	100
40) Benzene	4.64	78	15575	1.114	ug/l	98
52) Toluene	7.63	92	45060m	4.949	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF110518\
Data File : VF060586.D
Acq On : 5 Nov 2018 22:06
Operator : VA/AP
Sample : J5741-43
Misc : 6.98/5mL/MSVOA-F/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
TP-11

Manual Integrations
APPROVED

MMDadoda
11/6/2018 12:53:49 PM

Quant Time: Nov 06 07:28:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 03 05:29:31 2018
Response via : Initial Calibration

