

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF110518\
 Data File : VF060587.D
 Acq On : 5 Nov 2018 22:35
 Operator : VA/AP
 Sample : J5741-55RE
 Misc : 6.94/5mL/MSVOA-F/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 TP-14RE

Quant Time: Nov 06 07:31:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 03 05:29:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.87	168	100474	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.60	114	135139	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.78	117	88501	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.56	152	27549	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.84	65	17537	17.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	34.56%	
35) Dibromofluoromethane	4.11	113	25667	24.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	49.16%	
50) Toluene-d8	7.56	98	109597	35.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.88%	
62) 4-Bromofluorobenzene	11.41	95	29958	22.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	44.28%	
Target Compounds						
20) Methylene Chloride	2.19	84	5721	2.921	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF110518\
Data File : VF060587.D
Acq On : 5 Nov 2018 22:35
Operator : VA/AP
Sample : J5741-55RE
Misc : 6.94/5mL/MSVOA-F/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
TP-14RE

Quant Time: Nov 06 07:31:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 03 05:29:31 2018
Response via : Initial Calibration

