

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF110518\
 Data File : VF060597.D
 Acq On : 6 Nov 2018 6:14
 Operator : VA/AP
 Sample : J5767-07
 Misc : 5.27/5mL/MSVOA-F/SOIL
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 CL-02-110118-D

Quant Time: Nov 06 08:10:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 03 05:29:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.88	168	48354	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	5.60	114	63649	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.79	117	15566	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.56	152	10696	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.85	65	32828	67.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	134.38%	
35) Dibromofluoromethane	4.11	113	34457	70.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	140.14%	
50) Toluene-d8	7.57	98	5402	3.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	7.42%	
62) 4-Bromofluorobenzene	11.42	95	12471	19.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	39.14%	
Target Compounds						
20) Methylene Chloride	2.20	84	6393	14.892	ug/l #	79
95) Naphthalene	14.47	128	2986	6.552	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF110518\
Data File : VF060597.D
Acq On : 6 Nov 2018 6:14
Operator : VA/AP
Sample : J5767-07
Misc : 5.27/5mL/MSVOA-F/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
CL-02-110118-D

Quant Time: Nov 06 08:10:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 03 05:29:31 2018
Response via : Initial Calibration

