

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF110518\
 Data File : VF060598.D
 Acq On : 6 Nov 2018 6:43
 Operator : VA/AP
 Sample : J5776-01
 Misc : 5.33/5mL/MSVOA-F/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 NJ-GREEN

Manual Integrations
 APPROVED

MMDadoda
 11/6/2018 12:53:56 PM

Quant Time: Nov 06 08:14:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 03 05:29:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.88	168	195210	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	5.61	114	405365	50.00	ug/l	0.01
63) Chlorobenzene-d5	9.78	117	322884	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.55	152	158563	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.85	65	129361	65.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	131.18%	
35) Dibromofluoromethane	4.12	113	152497	48.69	ug/l	0.01
Spiked Amount 50.000			Recovery	=	97.38%	
50) Toluene-d8	7.56	98	313454	33.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	67.58%	
62) 4-Bromofluorobenzene	11.42	95	156902	38.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.32%	
Target Compounds						
20) Methylene Chloride	2.19	84	18382m	8.841	ug/l	Qvalue
30) Chloroform	3.87	83	6440	1.434	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF110518\
Data File : VF060598.D
Acq On : 6 Nov 2018 6:43
Operator : VA/AP
Sample : J5776-01
Misc : 5.33/5mL/MSVOA-F/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
NJ-GREEN

Manual Integrations
APPROVED

MMDadoda
11/6/2018 12:53:56 PM

Quant Time: Nov 06 08:14:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 03 05:29:31 2018
Response via : Initial Calibration

