

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF110518\
 Data File : VF060601.D
 Acq On : 6 Nov 2018 8:09
 Operator : VA/AP
 Sample : J5794-03
 Misc : 5.13/5mL/MSVOA-F/SOIL
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 TP-12B-S

Manual Integrations
 APPROVED

MMDadoda
 11/6/2018 12:53:59 PM

Quant Time: Nov 06 08:25:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 03 05:29:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.88	168	221998	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.61	114	434713	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.79	117	352484	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.56	152	144808	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.85	65	113692	50.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.38%	
35) Dibromofluoromethane	4.12	113	142002	42.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.56%	
50) Toluene-d8	7.56	98	345277	34.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.40%	
62) 4-Bromofluorobenzene	11.42	95	136558	31.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	62.76%	
Target Compounds						
8) Diethyl Ether	1.63	74	2414	3.887	ug/l	86
20) Methylene Chloride	2.18	84	18744m	7.293	ug/l	
80) 1,3,5-Trimethylbenzene	11.83	105	12144	1.305	ug/l	94
84) 1,2,4-Trimethylbenzene	12.22	105	19710	2.162	ug/l	97
95) Naphthalene	14.47	128	12537	2.032	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF110518\
Data File : VF060601.D
Acq On : 6 Nov 2018 8:09
Operator : VA/AP
Sample : J5794-03
Misc : 5.13/5mL/MSVOA-F/SOIL
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
TP-12B-S

Manual Integrations
APPROVED

MMDadoda
11/6/2018 12:53:59 PM

Quant Time: Nov 06 08:25:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 03 05:29:31 2018
Response via : Initial Calibration

