

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_F\DATA\VF110618\
 Data File : VF060623.D
 Acq On : 6 Nov 2018 20:18
 Operator : VA/AP
 Sample : J5819-04
 Misc : 4.6g/5mL/MSVOA-F/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 BES48_SB01_4-4.5

Quant Time: Nov 07 12:05:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 03 05:29:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.89	168	252083	50.00	ug/l	0.03
34) 1,4-Difluorobenzene	5.62	114	458265	50.00	ug/l	0.03
63) Chlorobenzene-d5	9.80	117	426968	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	12.57	152	212052	50.00	ug/l	0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.86	65	151532	59.49	ug/l	0.02
Spiked Amount 50.000			Recovery	=	118.98%	
35) Dibromofluoromethane	4.13	113	180185	50.89	ug/l	0.03
Spiked Amount 50.000			Recovery	=	101.78%	
50) Toluene-d8	7.58	98	445653	42.49	ug/l	0.02
Spiked Amount 50.000			Recovery	=	84.98%	
62) 4-Bromofluorobenzene	11.42	95	221064	48.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.38%	
Target Compounds						
52) Toluene	7.65	92	10240	1.28	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_F\DATA\VF110618\
Data File : VF060623.D
Acq On : 6 Nov 2018 20:18
Operator : VA/AP
Sample : J5819-04
Misc : 4.6g/5mL/MSVOA-F/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
BES48_SB01_4-4.5

Quant Time: Nov 07 12:05:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 03 05:29:31 2018
Response via : Initial Calibration

