

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF111218\
 Data File : VF060700.D
 Acq On : 12 Nov 2018 22:53
 Operator : VA/AP
 Sample : J5908-07RE
 Misc : 5.87g/5mL/MSVOA-F/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 SR-3-SRE

Manual Integrations
 APPROVED

MMDadoda
 11/13/2018 2:11:09 PM

Quant Time: Nov 13 04:56:12 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 03 05:29:31 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.85	168	230513	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	5.59	114	396279	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.77	117	287105	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	12.55	152	89293	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.83	65	135582	58.21	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	116.42%	
35) Dibromofluoromethane	4.10	113	149319	48.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.54%	
50) Toluene-d8	7.55	98	328983	36.27	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	72.54%	
62) 4-Bromofluorobenzene	11.40	95	115807	29.19	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	58.38%	
Target Compounds						
16) Acetone	2.23	43	7294	13.291	ug/l #	85
20) Methylene Chloride	2.17	84	12127m	2.232	ug/l	
68) m/p-Xylenes	10.12	106	7617	1.898	ug/l	76
69) o-Xylene	10.70	106	4630	1.056	ug/l	83
84) 1,2,4-Trimethylbenzene	12.20	105	10387	1.848	ug/l	92
95) Naphthalene	14.45	128	4680	1.230	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF111218\
Data File : VF060700.D
Acq On : 12 Nov 2018 22:53
Operator : VA/AP
Sample : J5908-07RE
Misc : 5.87g/5mL/MSVOA-F/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
SR-3-SRE

Manual Integrations
APPROVED

MMDadoda
11/13/2018 2:11:09 PM

Quant Time: Nov 13 04:56:12 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F110218S.M
Quant Title : SW846 8260
QLast Update : Sat Nov 03 05:29:31 2018
Response via : Initial Calibration

