

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_F\DATA\VF111418\
 Data File : VF060724.D
 Acq On : 14 Nov 2018 15:06
 Operator : VA/AP
 Sample : J5860-23
 Misc : 5.87g/5mL/MSVOA-F/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 EP2-SB-110(24-28)

Manual Integrations
 APPROVED

MMDadoda
 11/15/2018 1:30:42 PM

Quant Time: Nov 15 07:11:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F111318S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 14 03:11:14 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.85	168	229357	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	5.58	114	443933	50.00	ug/l	-0.04
63) Chlorobenzene-d5	9.76	117	404500	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	12.54	152	227064	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.82	65	129386	49.46	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	98.92%	
35) Dibromofluoromethane	4.09	113	148337	44.45	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	88.90%	
50) Toluene-d8	7.53	98	364899	42.03	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	84.06%	
62) 4-Bromofluorobenzene	11.40	95	185324	46.66	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	93.32%	
Target Compounds						
16) Acetone	2.23	43	19846	35.49	ug/l	Qvalue # 82
20) Methylene Chloride	2.22	84	18111m	7.30	ug/l	
30) Chloroform	3.83	83	32892	5.35	ug/l	97
52) Toluene	7.61	92	8598	1.12	ug/l	99
95) Naphthalene	14.45	128	9868	1.06	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_F\DATA\VF111418\
Data File : VF060724.D
Acq On : 14 Nov 2018 15:06
Operator : VA/AP
Sample : J5860-23
Misc : 5.87g/5mL/MSVOA-F/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
EP2-SB-110(24-28)

Manual Integrations
APPROVED

MMDadoda
11/15/2018 1:30:42 PM

Quant Time: Nov 15 07:11:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F111318S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 14 03:11:14 2018
Response via : Initial Calibration

