

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF111618\
 Data File : VF060766.D
 Acq On : 16 Nov 2018 16:01
 Operator : VA/AP
 Sample : J5767-15RE
 Misc : 5.74g/5mL/MSVOA-F/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 CL-01-111418-DRE

Quant Time: Nov 17 04:39:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F111318S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 14 03:11:14 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.89	168	90009	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.61	114	132446	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.79	117	116578	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.56	152	64285	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.86	65	25276	24.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	49.24%	
35) Dibromofluoromethane	4.13	113	32205	32.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	64.70%	
50) Toluene-d8	7.57	98	104026	39.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.96%	
62) 4-Bromofluorobenzene	11.42	95	47124	39.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.24%	
Target Compounds						
20) Methylene Chloride	2.20	84	2941	1.243	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF111618\
Data File : VF060766.D
Acq On : 16 Nov 2018 16:01
Operator : VA/AP
Sample : J5767-15RE
Misc : 5.74g/5mL/MSVOA-F/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
CL-01-111418-DRE

Quant Time: Nov 17 04:39:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F111318S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 14 03:11:14 2018
Response via : Initial Calibration

