

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF111618\
 Data File : VF060774.D
 Acq On : 16 Nov 2018 19:49
 Operator : VA/AP
 Sample : J5767-11RE
 Misc : 7.89g/5mL/MSVOA-F/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 CL-01-111418-BRE

Quant Time: Nov 17 04:24:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F111318S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 14 03:11:14 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	0.00	168	0	0.00	ug/l	-4.89
34) 1,4-Difluorobenzene	0.00	114	0	0.00	ug/l	-5.62
63) Chlorobenzene-d5	9.80	117	62	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.55	152	87	50.00	ug/l	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	4.70	65	64	0.00	ug/l	-0.16
Spiked Amount	50.000		Recovery	=	0.00%	
35) Dibromofluoromethane	4.12	113	63	0.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	0.00%	
50) Toluene-d8	0.00	98	0	0.00	ug/l	
Spiked Amount	50.000		Recovery	=	0.00%	
62) 4-Bromofluorobenzene	11.44	95	946	0.00	ug/l	0.03
Spiked Amount	50.000		Recovery	=	0.00%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
65) Chlorobenzene	9.77	112	67	53.552	ug/l #	42
67) Ethyl Benzene	9.89	91	59	25.018	ug/l #	45
73) Isopropylbenzene	11.10	105	416	59.404	ug/l #	48
74) N-amyl acetate	11.40	43	447	280.390	ug/l #	1
75) 1,1,2,2-Tetrachloroethane	11.69	83	738	637.423	ug/l #	25
76) 1,2,3-Trichloropropane	11.80	75	64	74.430	ug/l #	1
78) n-propylbenzene	11.60	91	1015	124.275	ug/l #	53
79) 2-Chlorotoluene	11.73	91	440	92.691	ug/l #	41
80) 1,3,5-Trimethylbenzene	11.88	105	427	76.803	ug/l #	27
81) trans-1,4-Dichloro-2-buten	11.80	75	64	151.501	ug/l #	40
82) 4-Chlorotoluene	11.93	91	411	84.516	ug/l #	43
83) tert-Butylbenzene	12.16	119	221	37.438	ug/l #	1
84) 1,2,4-Trimethylbenzene	12.37	105	169	29.392	ug/l #	78
85) sec-Butylbenzene	12.32	105	494	61.691	ug/l #	56
86) p-Isopropyltoluene	12.49	119	299	44.707	ug/l #	3
87) 1,3-Dichlorobenzene	12.64	146	60	20.683	ug/l #	1
88) 1,4-Dichlorobenzene	12.64	146	60	21.521	ug/l #	1
89) n-Butylbenzene	12.87	91	405	59.339	ug/l #	26
90) Hexachloroethane	12.94	117	77	47.072	ug/l #	21
91) 1,2-Dichlorobenzene	12.95	146	66	24.164	ug/l #	1
92) 1,2-Dibromo-3-Chloropropan	13.67	75	83	381.418	ug/l #	62
93) 1,2,4-Trichlorobenzene	14.11	180	81	41.260	ug/l #	1
95) Naphthalene	14.45	128	143	39.920	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF111618\
Data File : VF060774.D
Acq On : 16 Nov 2018 19:49
Operator : VA/AP
Sample : J5767-11RE
Misc : 7.89g/5mL/MSVOA-F/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
CL-01-111418-BRE

Quant Time: Nov 17 04:24:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F111318S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 14 03:11:14 2018
Response via : Initial Calibration

