

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF111918\
 Data File : VF060795.D
 Acq On : 20 Nov 2018 00:19
 Operator : VA/AP
 Sample : J5860-24
 Misc : 5.00µ/5mL/MSVOA-F/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 EP2-SB-111(14-16)

Quant Time: Nov 20 03:36:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F111318S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 14 03:11:14 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.88	168	199170	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	5.61	114	394216	50.00	ug/l	0.00
63) Chlorobenzene-d5	9.79	117	120186	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.56	152	158364	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.86	65	112088	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
35) Dibromofluoromethane	4.13	113	131388	44.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.68%	
50) Toluene-d8	7.56	98	109657	11.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	23.20%	
62) 4-Bromofluorobenzene	11.42	95	125013	34.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	68.78%	
Target Compounds						
16) Acetone	2.25	43	14093	29.023	ug/l #	88
20) Methylene Chloride	2.24	84	10940	4.156	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF111918\
Data File : VF060795.D
Acq On : 20 Nov 2018 00:19
Operator : VA/AP
Sample : J5860-24
Misc : 5.00µ/5mL/MSVOA-F/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
EP2-SB-111(14-16)

Quant Time: Nov 20 03:36:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F111318S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 14 03:11:14 2018
Response via : Initial Calibration

