


Data Path : Z:\voasrv\HPCHEM1\MSVOA_F\Data\VF112018\
 Data File : VF060801.D
 Acq On : 20 Nov 2018 12:34
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F111318S.M
 Title : SW846 8260
 Last Update : Wed Nov 14 03:11:14 2018



AutoFind: Scans 1093, 1094, 1095; Background Corrected with Scan 1087

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	26422	PASS
75	95	30	60	44.4	61061	PASS
95	95	100	100	100.0	137453	PASS
96	95	5	9	6.6	9005	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	72.4	99461	PASS
175	174	5	9	7.6	7527	PASS
176	174	95	101	97.0	96509	PASS
177	176	5	9	6.4	6148	PASS