

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF112918\
 Data File : VF060866.D
 Acq On : 29 Nov 2018 14:25
 Operator : VA/AP
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050

Quant Time: Nov 30 05:56:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 27 01:36:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.87	168	214580	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	5.60	114	375110	50.00	ug/l	-0.04
63) Chlorobenzene-d5	9.78	117	345897	50.00	ug/l	-0.03
72) 1,4-Dichlorobenzene-d4	12.56	152	168715	50.00	ug/l	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	4.84	65	139376	55.14	ug/l	-0.04
Spiked Amount			Recovery	=	110.28%	
35) Dibromofluoromethane	4.11	113	147197	49.46	ug/l	-0.04
Spiked Amount			Recovery	=	98.92%	
50) Toluene-d8	7.56	98	438260	50.00	ug/l	-0.04
Spiked Amount			Recovery	=	100.00%	
62) 4-Bromofluorobenzene	11.41	95	190658	47.86	ug/l	-0.03
Spiked Amount			Recovery	=	95.72%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	0.97	85	145949	45.489	ug/l	98
3) Chloromethane	1.10	50	91718	43.931	ug/l	95
4) Vinyl Chloride	1.14	62	91943	47.213	ug/l	97
5) Bromomethane	1.32	94	65557	47.740	ug/l	99
6) Chloroethane	1.37	64	39035	46.370	ug/l #	84
7) Trichlorofluoromethane	1.49	101	199264	47.689	ug/l	96
8) Diethyl Ether	1.63	74	29281	48.733	ug/l	90
9) 1,1,2-Trichlorotrifluoroet	1.82	101	95363	45.145	ug/l	93
10) Methyl Iodide	1.89	142	123612	38.499	ug/l	96
11) Tert butyl alcohol	2.57	59	28164	271.796	ug/l #	89
12) 1,1-Dichloroethene	1.78	96	78400	46.143	ug/l	99
13) Acrolein	2.00	56	23963	244.077	ug/l	95
14) Allyl chloride	2.10	41	106984	50.519	ug/l	97
15) Acrylonitrile	2.94	53	56296	242.330	ug/l	93
16) Acetone	2.24	43	137628	295.247	ug/l	98
17) Carbon Disulfide	1.82	76	204459	44.880	ug/l	98
18) Methyl Acetate	2.35	43	44237	56.790	ug/l	94
19) Methyl tert-butyl Ether	2.44	73	191708	50.097	ug/l	99
20) Methylene Chloride	2.24	84	48510	28.367	ug/l	93
21) trans-1,2-Dichloroethene	2.31	96	82846	50.577	ug/l	94
22) Diisopropyl ether	2.80	45	300545	52.138	ug/l	99
23) Vinyl Acetate	3.19	43	719483	273.051	ug/l	98
24) 1,1-Dichloroethane	2.87	63	169081	49.300	ug/l	98
25) 2-Butanone	4.33	43	220866	273.936	ug/l	97
26) 2,2-Dichloropropane	3.60	77	106554	47.923	ug/l	97
27) cis-1,2-Dichloroethene	3.48	96	134019	50.256	ug/l	91
28) Bromochloromethane	3.72	49	84108	51.100	ug/l	92
29) Tetrahydrofuran	4.07	42	87797	256.893	ug/l	97
30) Chloroform	3.87	83	252562	47.848	ug/l	96
31) Cyclohexane	3.70	56	164081	46.099	ug/l	92
32) 1,1,1-Trichloroethane	4.09	97	172038	46.996	ug/l	96
36) 1,1-Dichloropropene	4.28	75	184194	44.271	ug/l	96
37) Ethyl Acetate	4.11	43	81571	48.245	ug/l	96
38) Carbon Tetrachloride	3.99	117	157735	42.923	ug/l	97

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF112918\
 Data File : VF060866.D
 Acq On : 29 Nov 2018 14:25
 Operator : VA/AP
 Sample : VSTDCCC050
 Misc : 5.00g/5mL/MSVOA-F/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050

Quant Time: Nov 30 05:56:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 27 01:36:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	5.45	83	183782	42.150	ug/l	98
40) Benzene	4.64	78	409687	44.572	ug/l	99
41) Methacrylonitrile	4.75	41	41100	46.694	ug/l #	87
42) 1,2-Dichloroethane	4.95	62	154339	45.994	ug/l	100
43) Isopropyl Acetate	6.96	43	107718	51.731	ug/l #	98
44) Trichloroethene	5.50	130	127129	42.301	ug/l	96
45) 1,2-Dichloropropane	6.24	63	108457	45.433	ug/l	95
46) Dibromomethane	6.09	93	77441	47.821	ug/l	98
47) Bromodichloromethane	6.39	83	199115	46.294	ug/l	99
48) Methyl methacrylate	6.72	41	70208	45.875	ug/l	96
49) 1,4-Dioxane	6.71	88	11145	995.024	ug/l #	77
51) 4-Methyl-2-Pentanone	8.27	43	366171	234.853	ug/l	98
52) Toluene	7.63	92	273720	42.396	ug/l	99
53) t-1,3-Dichloropropene	8.29	75	156965	44.866	ug/l	98
54) cis-1,3-Dichloropropene	7.31	75	199396	46.038	ug/l	98
55) 1,1,2-Trichloroethane	8.49	97	87988	46.770	ug/l	97
56) Ethyl methacrylate	8.63	69	101130	48.391	ug/l	95
57) 1,3-Dichloropropane	8.86	76	159946	48.595	ug/l	96
58) 2-Chloroethyl Vinyl ether	7.30	63	50186	234.773	ug/l	97
59) 2-Hexanone	9.50	43	275382	251.758	ug/l	97
60) Dibromochloromethane	8.72	129	126087	44.982	ug/l	92
61) 1,2-Dibromoethane	9.00	107	91692	45.117	ug/l	92
64) Tetrachloroethene	8.16	164	113910	43.116	ug/l	96
65) Chlorobenzene	9.80	112	303234	45.064	ug/l	98
66) 1,1,1,2-Tetrachloroethane	9.94	131	127610	44.113	ug/l	93
67) Ethyl Benzene	9.90	91	567797	45.324	ug/l	99
68) m/p-Xylenes	10.14	106	388520	86.230	ug/l	97
69) o-Xylene	10.72	106	220530	44.387	ug/l	95
70) Styrene	10.79	104	293006	42.543	ug/l	98
71) Bromoform	10.78	173	65343	46.764	ug/l #	95
73) Isopropylbenzene	11.13	105	661721	49.631	ug/l	97
74) N-amyl acetate	11.37	43	165905	53.887	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.70	83	113939	51.925	ug/l	100
76) 1,2,3-Trichloropropane	11.80	75	80907	50.727	ug/l	94
77) Bromobenzene	11.50	156	141327	48.349	ug/l	85
78) n-propylbenzene	11.60	91	719077	46.054	ug/l	97
79) 2-Chlorotoluene	11.73	91	425624	47.541	ug/l	96
80) 1,3,5-Trimethylbenzene	11.83	105	508967	47.023	ug/l	95
81) trans-1,4-Dichloro-2-buten	11.88	75	51240	65.569	ug/l	93
82) 4-Chlorotoluene	11.91	91	442868	47.006	ug/l	96
83) tert-Butylbenzene	12.13	119	528772	47.959	ug/l	99
84) 1,2,4-Trimethylbenzene	12.21	105	521789	48.303	ug/l	99
85) sec-Butylbenzene	12.31	105	704785	46.696	ug/l	99
86) p-Isopropyltoluene	12.47	119	558358	44.125	ug/l	99
87) 1,3-Dichlorobenzene	12.48	146	248004	44.604	ug/l	98
88) 1,4-Dichlorobenzene	12.57	146	246516	45.495	ug/l	96
89) n-Butylbenzene	12.85	91	613886	47.978	ug/l	97
90) Hexachloroethane	12.92	117	143194	46.993	ug/l	95
91) 1,2-Dichlorobenzene	12.95	146	247841	47.070	ug/l	98
92) 1,2-Dibromo-3-Chloropropan	13.65	75	22111	52.119	ug/l	94


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF112918\
Data File : VF060866.D
Acq On : 29 Nov 2018 14:25
Operator : VA/AP
Sample : VSTDCCC050
Misc : 5.00g/5mL/MSVOA-F/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Quant Time: Nov 30 05:56:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 27 01:36:26 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	14.20	180	178394	48.489	ug/l	97
94) Hexachlorobutadiene	14.19	225	111520	45.472	ug/l	99
95) Naphthalene	14.46	128	362625	50.841	ug/l	99
96) 1,2,3-Trichlorobenzene	14.61	180	167591	49.060	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 4.87 min Scan# 430

Delta R.T. -0.04 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion:168 Resp: 214580

Ion Ratio Lower Upper

168 100

99 71.2 52.6 79.0



Abundance Ion 168.00 (167.70 to 168.70): VF0

Ion 99.00 (98.70 to 99.70): VF0

4.87

60000

40000



20000

0

Time-->

4.80

5.00



#2

Dichlorodifluoromethane

Concen: 45.489 ug/l

RT: 0.97 min Scan# 34

Delta R.T. -0.01 min

Lab File: VF060866.D


Acq: 29 Nov 2018 14:25

Tgt Ion: 85 Resp: 145949

Ion Ratio Lower Upper

85 100

87 32.1 16.7 50.0



Abundance Ion 85.00 (84.70 to 85.70): VF0

Ion 87.00 (86.70 to 87.70): VF0

0.97

50000

40000

30000

20000

10000

0

Time-->


0.80


0.90

1.00

1.10

1.20





#3

Chloromethane

Concen: 43.931 ug/l

RT: 1.10 min Scan# 48

Delta R.T. 0.00 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion: 50 Resp: 91718

Ion Ratio Lower Upper




50 100

52 36.6 27.0 40.4



Abundance Ion 50.00 (49.70 to 50.70): VF0

Ion 52.00 (51.70 to 52.70): VF0



#4

Vinyl Chloride

Concen: 47.213 ug/l

RT: 1.14 min Scan# 52

Delta R.T. -0.01 min

Lab File: VF060866.D


Acq: 29 Nov 2018 14:25

Tgt Ion: 62 Resp: 91943

Ion Ratio Lower Upper



62 100


64 31.1 26.4 39.6



Abundance Ion 62.00 (61.70 to 62.70): VF0

Ion 64.00 (63.70 to 64.70): VF0





#5

Bromomethane

Concen: 47.740 ug/l

RT: 1.32 min Scan# 70

Delta R.T. -0.02 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion: 94 Resp: 65557

Ion Ratio Lower Upper




94 100

96 93.5 75.9 113.9



Abundance Ion 94.00 (93.70 to 94.70): VF0

Ion 96.00 (95.70 to 96.70): VF0



#6

Chloroethane

Concen: 46.370 ug/l

RT: 1.37 min Scan# 75

Delta R.T. -0.02 min

Lab File: VF060866.D


Acq: 29 Nov 2018 14:25

Tgt Ion: 64 Resp: 39035

Ion Ratio Lower Upper



64 100


66 26.0 28.1 42.1#



Abundance Ion 64.00 (63.70 to 64.70): VF0

Ion 66.00 (65.70 to 66.70): VF0









#7

Trichlorofluoromethane
 Concen: 47.689 ug/l
 RT: 1.49 min Scan# 87
 Delta R.T. -0.01 min
 Lab File: VF060866.D
 Acq: 29 Nov 2018 14:25

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050




Tot Ion:101 Resp: 199264
 Ion Ratio Lower Upper
 101 100
 103 65.6 55.4 83.0




#8

Diethyl Ether
 Concen: 48.733 ug/l
 RT: 1.63 min Scan# 101
 Delta R.T. -0.02 min
 Lab File: VF060866.D
 Acq: 29 Nov 2018 14:25

Tgt Ion: 74 Resp: 29281
 Ion Ratio Lower Upper
 74 100
 45 153.6 71.0 212.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 45.145 ug/l

RT: 1.82 min Scan# 121

Delta R.T. -0.01 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion:101 Resp: 95363

Ion Ratio Lower Upper

101 100

85 49.6 42.8 64.2




151 61.0 54.4 81.6



Abundance Ion 101.00 (100.70 to 101.70): V

Ion 85.00 (84.70 to 85.70): VF0

Ion 151.00 (150.70 to 151.70): V



#10

Methyl Iodide

Concen: 38.499 ug/l

RT: 1.89 min Scan# 128

Delta R.T. -0.02 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25


Tgt Ion:142 Resp: 123612

Ion Ratio Lower Upper

142 100

127 68.1 51.9 77.9



141 14.7 12.7 19.1




Abundance Ion 142.00 (141.70 to 142.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 141.00 (140.70 to 141.70): V









#11

Tert butyl alcohol
Concen: 271.796 ug/l
RT: 2.57 min Scan# 197
Delta R.T. -0.01 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050




Tot Ion: 59 Resp: 28164
Ion Ratio Lower Upper
59 100
57 11.1 12.7 19.1#




#12

1,1-Dichloroethene
Concen: 46.143 ug/l
RT: 1.78 min Scan# 117
Delta R.T. -0.02 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25

Tgt Ion: 96 Resp: 78400
Ion Ratio Lower Upper
96 100
61 220.5 177.3 265.9
98 61.5 49.6 74.4





#13

Acrolein

Concen: 244.077 ug/l

RT: 2.00 min Scan# 139

Delta R.T. -0.02 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion: 56 Resp: 23963

Ion Ratio Lower Upper

56 100

55 80.6 61.2 91.8



#14

Allyl chloride

Concen: 50.519 ug/l

RT: 2.10 min Scan# 149

Delta R.T. -0.02 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25




Tgt Ion: 41 Resp: 106984


Ion Ratio Lower Upper

41 100

39 111.6 91.7 137.5

76 30.3 23.5 35.3





#15

Acrylonitrile

Concen: 242.330 ug/l

RT: 2.94 min Scan# 234

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion: 53 Resp: 56296

Ion Ratio Lower Upper

53 100

52 86.1 76.5 114.7




51 48.7 40.7 61.1



Abundance Ion 53.00 (52.70 to 53.70): VF0

Ion 52.00 (51.70 to 52.70): VF0

Ion 51.00 (50.70 to 51.70): VF0



#16

Acetone

Concen: 295.247 ug/l

RT: 2.24 min Scan# 163

Delta R.T. -0.02 min

Lab File: VF060866.D


Acq: 29 Nov 2018 14:25

Tgt Ion: 43 Resp: 137628

Ion Ratio Lower Upper



43 100


58 13.7 11.4 17.2



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0









#17
Carbon Disulfide
Concen: 44.880 ug/l
RT: 1.82 min Scan# 121
Delta R.T. -0.03 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 76 Resp: 204459
Ion Ratio Lower Upper
76 100
78 12.0 10.2 15.2



#18
Methyl Acetate
Concen: 56.790 ug/l
RT: 2.35 min Scan# 174
Delta R.T. -0.02 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25

Tgt Ion: 43 Resp: 44237
Ion Ratio Lower Upper
43 100
74 13.8 13.3 19.9









#19
Methyl tert-butyl Ether
Concen: 50.097 ug/l
RT: 2.44 min Scan# 183
Delta R.T. -0.02 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 73 Resp: 191708
Ion Ratio Lower Upper
73 100
57 18.7 15.3 22.9



#20
Methylene Chloride
Concen: 28.367 ug/l
RT: 2.24 min Scan# 163
Delta R.T. -0.02 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25

Tgt Ion: 84 Resp: 48510
Ion Ratio Lower Upper
84 100
49 149.7 111.3 166.9
51 44.2 38.0 57.0
86 66.9 49.8 74.8





#21

trans-1,2-Dichloroethene

Concen: 50.577 ug/l

RT: 2.31 min Scan# 170

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050





Tot Ion: 96 Resp: 82846

Ion Ratio Lower Upper

96 100

61 175.5 134.5 201.7

98 58.7 51.6 77.4



#22

Diisopropyl ether

Concen: 52.138 ug/l

RT: 2.80 min Scan# 220

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Tgt Ion: 45 Resp: 300545




Ion Ratio Lower Upper


45 100

43 50.4 41.0 61.6

87 18.2 14.6 22.0

59 9.0 7.5 11.3





#23

Vinyl Acetate

Concen: 273.051 ug/l

RT: 3.19 min Scan# 259

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion: 43 Resp: 719483

Ion Ratio Lower Upper

43 100

86 7.3 5.4 8.0



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 86.00 (85.70 to 86.70): VF0

3.19

250000

200000



150000

100000

50000

0

Time-->



#24

1,1-Dichloroethane

Concen: 49.300 ug/l

RT: 2.87 min Scan# 227

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25


Tgt Ion: 63 Resp: 169081

Ion Ratio Lower Upper

63 100

98 4.5 2.7 8.1

100 2.9 1.5 4.5



Abundance Ion 63.00 (62.70 to 63.70): VF0

Ion 98.00 (97.70 to 98.70): VF0

Ion 100.00 (99.70 to 100.70): VF0

2.87

80000


60000


40000

20000

0

Time-->





#25

2-Butanone

Concen: 273.936 ug/l

RT: 4.33 min Scan# 375

Delta R.T. -0.04 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion: 43 Resp: 220866

Ion Ratio Lower Upper




43 100

72 18.0 15.4 23.2



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 72.00 (71.70 to 72.70): VF0



#26

2,2-Dichloropropane

Concen: 47.923 ug/l

RT: 3.60 min Scan# 301

Delta R.T. -0.04 min

Lab File: VF060866.D


Acq: 29 Nov 2018 14:25

Tgt Ion: 77 Resp: 106554

Ion Ratio Lower Upper



77 100


97 18.4 8.6 25.8



Abundance Ion 77.00 (76.70 to 77.70): VF0

Ion 97.00 (96.70 to 97.70): VF0






#27




cis-1,2-Dichloroethene
 Concen: 50.256 ug/l
 RT: 3.48 min Scan# 289
 Delta R.T. -0.03 min
 Lab File: VF060866.D
 Acq: 29 Nov 2018 14:25

Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050

Tot Ion	Ratio	Lower	Upper
96	100		
61	134.1	0.0	290.6
98	62.1	0.0	137.8




Abundance Ion 96.00 (95.70 to 96.70): VF0
 Ion 61.00 (60.70 to 61.70): VF0
 Ion 98.00 (97.70 to 98.70): VF0





#28


Bromochloromethane
 Concen: 51.100 ug/l
 RT: 3.72 min Scan# 313
 Delta R.T. -0.04 min
 Lab File: VF060866.D
 Acq: 29 Nov 2018 14:25

Tgt Ion	Ratio	Lower	Upper
49	100		
129	0.0	0.0	4.4
130	73.4	64.2	96.4



Abundance Ion 49.00 (48.70 to 49.70): VF0
 Ion 129.00 (128.70 to 129.70): V
 Ion 130.00 (129.70 to 130.70): V





#29

Tetrahydrofuran

Concen: 256.893 ug/l

RT: 4.07 min Scan# 349

Delta R.T. -0.04 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion: 42 Resp: 87797

Ion Ratio Lower Upper




42 100

72 38.2 31.9 47.9

71 37.4 28.7 43.1



Abundance Ion 42.00 (41.70 to 42.70): VF0
Ion 72.00 (71.70 to 72.70): VF0
Ion 71.00 (70.70 to 71.70): VF0



#30

Chloroform

Concen: 47.848 ug/l

RT: 3.87 min Scan# 328

Delta R.T. -0.03 min

Lab File: VF060866.D


Acq: 29 Nov 2018 14:25

Tgt Ion: 83 Resp: 252562



Ion Ratio Lower Upper


83 100

85 63.9 53.4 80.2



Abundance Ion 83.00 (82.70 to 83.70): VF0
Ion 85.00 (84.70 to 85.70): VF0





#31

Cyclohexane

Concen: 46.099 ug/l

RT: 3.70 min Scan# 311

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion: 56 Resp: 164081

Ion Ratio Lower Upper




56 100

69 33.1 23.0 34.4

84 95.2 70.4 105.6



Abundance Ion 56.00 (55.70 to 56.70): VF0
Ion 69.00 (68.70 to 69.70): VF0
Ion 84.00 (83.70 to 84.70): VF0



#32

1,1,1-Trichloroethane

Concen: 46.996 ug/l

RT: 4.09 min Scan# 351

Delta R.T. -0.04 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25


Tgt Ion: 97 Resp: 172038

Ion Ratio Lower Upper



97 100


99 61.6 53.5 80.3

61 48.2 38.6 58.0



Abundance Ion 97.00 (96.70 to 97.70): VF0
Ion 99.00 (98.70 to 99.70): VF0
Ion 61.00 (60.70 to 61.70): VF0





#33

1,2-Dichloroethane-d4

Concen: 55.137 ug/l

RT: 4.84 min Scan# 427

Delta R.T. -0.04 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion: 65 Resp: 139376

Ion Ratio Lower Upper




65 100

67 48.6 0.0 96.4



Abundance Ion 65.00 (64.70 to 65.70): VF0

Ion 67.00 (66.70 to 67.70): VF0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 5.60 min Scan# 504

Delta R.T. -0.04 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25


Tgt Ion: 114 Resp: 375110

Ion Ratio Lower Upper

114 100

63 22.1 0.0 41.6



88 20.2 0.0 36.6




Abundance Ion 114.00 (113.70 to 114.70): V

Ion 63.00 (62.70 to 63.70): VF0

Ion 88.00 (87.70 to 88.70): VF0









#35
 Dibromofluoromethane
 Concen: 49.464 ug/l
 RT: 4.11 min Scan# 353
 Delta R.T. -0.04 min
 Lab File: VF060866.D
 Acq: 29 Nov 2018 14:25




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050


Tot Ion:113 Resp: 147197
 Ion Ratio Lower Upper
 113 100
 111 103.4 83.0 124.6
 192 15.2 12.3 18.5



#36
 1,1-Dichloropropene
 Concen: 44.271 ug/l
 RT: 4.28 min Scan# 370
 Delta R.T. -0.04 min
 Lab File: VF060866.D
 Acq: 29 Nov 2018 14:25

Tgt Ion: 75 Resp: 184194
 Ion Ratio Lower Upper
 75 100
 110 28.8 15.6 46.8
 77 31.3 23.4 35.0









#37
Ethyl Acetate
Concen: 48.245 ug/l
RT: 4.11 min Scan# 353
Delta R.T. -0.04 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Tot Ion: 43 Resp: 81571
Ion Ratio Lower Upper
43 100
61 89.5 68.6 103.0
70 9.9 9.0 13.4




Abundance Ion 43.00 (42.70 to 43.70): VF0
Ion 61.00 (60.70 to 61.70): VF0
Ion 70.00 (69.70 to 70.70): VF0






#38
Carbon Tetrachloride
Concen: 42.923 ug/l
RT: 3.99 min Scan# 341
Delta R.T. -0.03 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25

Tgt Ion: 117 Resp: 157735
Ion Ratio Lower Upper
117 100
119 90.1 74.4 111.6
121 32.0 23.5 35.3



Abundance Ion 117.00 (116.70 to 117.70): V
Ion 119.00 (118.70 to 119.70): V
Ion 121.00 (120.70 to 121.70): V









#39
Methvlcvclohexane
Concen: 42.150 ug/l
RT: 5.45 min Scan# 489
Delta R.T. -0.04 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 83 Resp: 183782
Ion Ratio Lower Upper
83 100
55 88.9 72.1 108.1
98 46.5 39.1 58.7



#40
Benzene
Concen: 44.572 ug/l
RT: 4.64 min Scan# 406
Delta R.T. -0.04 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25

Tgt Ion: 78 Resp: 409687
Ion Ratio Lower Upper
78 100
77 23.7 18.7 28.1









#41
Methacrylonitrile
Concen: 46.694 ug/l
RT: 4.75 min Scan# 418
Delta R.T. -0.03 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion:	41	Resp:	41100
Ion	Ratio	Lower	Upper
41	100		
39	52.9	53.0	79.6#
67	47.9	43.8	65.6
52	40.2	37.8	56.8



#42
1,2-Dichloroethane
Concen: 45.994 ug/l
RT: 4.95 min Scan# 438
Delta R.T. -0.03 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25

Tgt Ion:	62	Resp:	154339
Ion	Ratio	Lower	Upper
62	100		
98	8.6	0.0	17.2





#43

Isopropyl Acetate

Concen: 51.731 ug/l

RT: 6.96 min Scan# 642

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion: 43 Resp: 107718

Ion Ratio Lower Upper

43 100

61 27.9 21.4 32.0




87 0.4 0.2 0.4#



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 61.00 (60.70 to 61.70): VF0

Ion 87.00 (86.70 to 87.70): VF0



#44

Trichloroethene

Concen: 42.301 ug/l

RT: 5.50 min Scan# 494

Delta R.T. -0.04 min

Lab File: VF060866.D


Acq: 29 Nov 2018 14:25

Tgt Ion: 130 Resp: 127129

Ion Ratio Lower Upper



130 100


95 102.2 0.0 195.6



Abundance Ion 129.90 (129.60 to 130.60): V

Ion 94.90 (94.60 to 95.60): VF0





#45

1,2-Dichloropropane

Concen: 45.433 ug/l

RT: 6.24 min Scan# 569

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion: 63 Resp: 108457

Ion Ratio Lower Upper

63 100

65 31.5 23.2 34.8



#46

Dibromomethane

Concen: 47.821 ug/l

RT: 6.09 min Scan# 553

Delta R.T. -0.04 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25




Tgt Ion: 93 Resp: 77441


Ion Ratio Lower Upper

93 100

95 81.7 64.7 97.1

174 88.6 73.0 109.6





#47

Bromodichloromethane

Concen: 46.294 ug/l

RT: 6.39 min Scan# 584

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion: 83 Resp: 199115

Ion Ratio Lower Upper

83 100

85 65.0 51.1 76.7




127 8.8 6.8 10.2



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 85.00 (84.70 to 85.70): VF0

Ion 127.00 (126.70 to 127.70): V



#48

Methyl methacrylate

Concen: 45.875 ug/l

RT: 6.72 min Scan# 617

Delta R.T. -0.04 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25


Tgt Ion: 41 Resp: 70208

Ion Ratio Lower Upper

41 100

69 70.6 53.9 80.9



39 65.1 49.0 73.6




Abundance Ion 41.00 (40.70 to 41.70): VF0

Ion 69.00 (68.70 to 69.70): VF0

Ion 39.00 (38.70 to 39.70): VF0





#49

1,4-Dioxane

Concen: 995.024 ug/l

RT: 6.71 min Scan# 616

Delta R.T. -0.04 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050





Tot Ion: 88 Resp: 11145

Ion Ratio Lower Upper

88 100

43 62.0 42.0 63.0

58 51.4 61.8 92.8#



#50

Toluene-d8

Concen: 49.996 ug/l

RT: 7.56 min Scan# 702

Delta R.T. -0.04 min

Lab File: VF060866.D




Acq: 29 Nov 2018 14:25


Tgt Ion: 98 Resp: 438260

Ion Ratio Lower Upper

98 100

100 68.5 59.3 88.9





#51

4-Methyl-2-Pentanone

Concen: 234.853 ug/l

RT: 8.27 min Scan# 774

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion: 43 Resp: 366171

Ion Ratio Lower Upper




43 100

58 38.3 29.7 44.5



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 58.00 (57.70 to 58.70): VF0



#52

Toluene

Concen: 42.396 ug/l

RT: 7.63 min Scan# 710

Delta R.T. -0.03 min

Lab File: VF060866.D


Acq: 29 Nov 2018 14:25

Tgt Ion: 92 Resp: 273720

Ion Ratio Lower Upper



92 100


91 159.4 126.2 189.2



Abundance Ion 92.00 (91.70 to 92.70): VF0

Ion 91.00 (90.70 to 91.70): VF0





#53

t-1,3-Dichloropropene

Concen: 44.866 ug/l

RT: 8.29 min Scan# 776

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion: 75 Resp: 156965

Ion Ratio Lower Upper

75 100

77 32.9 27.4 41.2



#54

cis-1,3-Dichloropropene

Concen: 46.038 ug/l

RT: 7.31 min Scan# 677

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25




Tgt Ion: 75 Resp: 199396


Ion Ratio Lower Upper

75 100

77 34.2 25.0 37.6

39 63.4 51.1 76.7





#55

1,1,2-Trichloroethane

Concen: 46.770 ug/l

RT: 8.49 min Scan# 797

Delta R.T. -0.04 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050

Tot Ion: 97 Resp: 87988


Ion Ratio Lower Upper

97 100

83 86.6 67.8 101.6

85 57.2 45.0 67.6

99 67.2 50.5 75.7






Abundance Ion 97.00 (96.70 to 97.70): VF0

Ion 82.95 (82.65 to 83.65): VF0

Ion 84.95 (84.65 to 85.65): VF0

Ion 98.95 (98.65 to 99.65): VF0



#56

Ethyl methacrylate

Concen: 48.391 ug/l

RT: 8.63 min Scan# 811

Delta R.T. -0.02 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25


Tgt Ion: 69 Resp: 101130

Ion Ratio Lower Upper

69 100

41 74.2 62.6 94.0



39 43.9 37.8 56.6




Abundance Ion 69.00 (68.70 to 69.70): VF0

Ion 41.00 (40.70 to 41.70): VF0

Ion 39.00 (38.70 to 39.70): VF0





#57

1,3-Dichloropropane

Concen: 48.595 ug/l

RT: 8.86 min Scan# 834

Delta R.T. -0.04 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion: 76 Resp: 159946

Ion Ratio Lower Upper

76 100

78 31.7 27.0 40.4



#58

2-Chloroethyl Vinyl ether

Concen: 234.773 ug/l

RT: 7.30 min Scan# 676

Delta R.T. -0.04 min

Lab File: VF060866.D




Acq: 29 Nov 2018 14:25


Tgt Ion: 63 Resp: 50186

Ion Ratio Lower Upper

63 100

106 27.0 23.0 34.6









#59
2-Hexanone
Concen: 251.758 ug/l
RT: 9.50 min Scan# 899
Delta R.T. -0.03 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 43 Resp: 275382
Ion Ratio Lower Upper
43 100
58 53.1 25.6 76.6



#60
Dibromochloromethane
Concen: 44.982 ug/l
RT: 8.72 min Scan# 820
Delta R.T. -0.03 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25

Tgt Ion: 129 Resp: 126087
Ion Ratio Lower Upper
129 100
127 81.3 37.3 111.9





#61

1,2-Dibromoethane

Concen: 45.117 ug/l

RT: 9.00 min Scan# 848

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion:107 Resp: 91692

Ion Ratio Lower Upper




107 100

109 99.5 73.6 110.4



Abundance Ion 107.00 (106.70 to 107.70): V

Ion 109.00 (108.70 to 109.70): V



#62

4-Bromofluorobenzene

Concen: 47.856 ug/l

RT: 11.41 min Scan# 1093

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25


Tgt Ion: 95 Resp: 190658

Ion Ratio Lower Upper

95 100

174 63.3 0.0 138.2



176 61.7 0.0 146.2




Abundance Ion 95.00 (94.70 to 95.70): VF0

Ion 174.00 (173.70 to 174.70): V

Ion 176.00 (175.70 to 176.70): V





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 9.78 min Scan# 928

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050



Tot Ion:117 Resp: 345897

Ion Ratio Lower Upper

117 100

82 58.5 46.6 69.8



119 33.7 26.4 39.6



Abundance Ion 117.00 (116.70 to 117.70): V

Ion 82.00 (81.70 to 82.70): VF0

Ion 119.00 (118.70 to 119.70): V



#64

Tetrachloroethene

Concen: 43.116 ug/l

RT: 8.16 min Scan# 763

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Tgt Ion:164 Resp: 113910



Ion Ratio Lower Upper

164 100

166 132.1 100.3 150.5

129 99.2 77.9 116.9

131 93.4 77.5 116.3





Abundance Ion 163.85 (163.55 to 164.55): V

Ion 165.80 (165.50 to 166.50): V

Ion 128.90 (128.60 to 129.60): V

Ion 130.90 (130.60 to 131.60): V









#65
Chlorobenzene
Concen: 45.064 ug/l
RT: 9.80 min Scan# 930
Delta R.T. -0.03 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion:112 Resp: 303234
Ion Ratio Lower Upper
112 100
114 33.6 26.0 39.0



#66
1,1,1,2-Tetrachloroethane
Concen: 44.113 ug/l
RT: 9.94 min Scan# 944
Delta R.T. -0.03 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25

Tgt Ion:131 Resp: 127610
Ion Ratio Lower Upper
131 100
133 101.7 46.5 139.3
119 65.8 31.1 93.3





#67

Ethyl Benzene

Concen: 45.324 ug/l

RT: 9.90 min Scan# 940

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion: 91 Resp: 567797

Ion Ratio Lower Upper

91 100

106 28.3 22.2 33.2



#68

m/p-Xylenes

Concen: 86.230 ug/l

RT: 10.14 min Scan# 964

Delta R.T. -0.03 min

Lab File: VF060866.D




Acq: 29 Nov 2018 14:25


Tgt Ion: 106 Resp: 388520

Ion Ratio Lower Upper

106 100

91 203.6 166.7 250.1









#69
o-Xylene
Concen: 44.387 ug/l
RT: 10.72 min Scan# 1023
Delta R.T. -0.03 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion:106 Resp: 220530
Ion Ratio Lower Upper
106 100
91 217.9 104.8 314.6



#70
Styrene
Concen: 42.543 ug/l
RT: 10.79 min Scan# 1030
Delta R.T. -0.03 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25

Tgt Ion:104 Resp: 293006
Ion Ratio Lower Upper
104 100
78 48.2 37.3 55.9
103 44.9 36.8 55.2





#71

Bromoform

Concen: 46.764 ug/l

RT: 10.78 min Scan# 1029

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion:173 Resp: 65343

Ion Ratio Lower Upper

173 100

175 50.3 25.5 76.5




254 0.0 7.5 11.3#



Abundance Ion 173.00 (172.70 to 173.70): V

Ion 175.00 (174.70 to 175.70): V

Ion 254.00 (253.70 to 254.70): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.56 min Scan# 1209

Delta R.T. -0.02 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25


Tgt Ion:152 Resp: 168715

Ion Ratio Lower Upper

152 100

115 56.4 30.1 90.5



150 159.1 0.0 327.6




Abundance Ion 152.00 (151.70 to 152.70): V

Ion 115.00 (114.70 to 115.70): V

Ion 150.00 (149.70 to 150.70): V





#73

Isopropylbenzene

Concen: 49.631 ug/l

RT: 11.13 min Scan# 1064

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :


VSTDCCC050

Tot Ion:105 Resp: 661721

Ion Ratio Lower Upper

105 100

120 25.8 12.2 36.4



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V

11.13

400000



300000

200000

100000

0

Time--> 11.00 11.20



#74

N-ethyl acetate

Concen: 53.887 ug/l

RT: 11.37 min Scan# 1089

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Tgt Ion: 43 Resp: 165905


Ion Ratio Lower Upper

43 100

70 34.6 29.6 44.4

55 20.9 17.2 25.8

61 25.2 21.8 32.6



Abundance Ion 43.00 (42.70 to 43.70): VF0

Ion 70.00 (69.70 to 70.70): VF0

Ion 55.00 (54.70 to 55.70): VF0

Ion 61.00 (60.70 to 61.70): VF0


150000


100000

50000

0

Time--> 11.20 11.40





#75

1,1,2,2-Tetrachloroethane

Concen: 51.925 ug/l

RT: 11.70 min Scan# 1122

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion: 83 Resp: 113939

Ion Ratio Lower Upper

83 100

131 8.6 4.7 14.1




85 65.1 32.6 97.8



Abundance Ion 83.00 (82.70 to 83.70): VF0

Ion 131.00 (130.70 to 131.70): V

Ion 85.00 (84.70 to 85.70): VF0



#76

1,2,3-Trichloropropane

Concen: 50.727 ug/l

RT: 11.80 min Scan# 1132

Delta R.T. -0.03 min

Lab File: VF060866.D


Acq: 29 Nov 2018 14:25

Tgt Ion: 75 Resp: 80907

Ion Ratio Lower Upper



75 100


77 33.3 18.4 55.4



Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 77.00 (76.70 to 77.70): VF0





#77

Bromobenzene

Concen: 48.349 ug/l

RT: 11.50 min Scan# 1102

Delta R.T. -0.03 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion:156 Resp: 141327

Ion Ratio Lower Upper

156 100

77 168.1 69.7 209.0




158 97.7 47.8 143.4



Abundance Ion 156.00 (155.70 to 156.70): V

Ion 77.00 (76.70 to 77.70): VF0

Ion 158.00 (157.70 to 158.70): V



#78

n-propylbenzene

Concen: 46.054 ug/l

RT: 11.60 min Scan# 1112

Delta R.T. -0.02 min

Lab File: VF060866.D


Acq: 29 Nov 2018 14:25

Tgt Ion: 91 Resp: 719077

Ion Ratio Lower Upper



91 100


120 21.9 10.3 30.8



Abundance Ion 91.00 (90.70 to 91.70): VF0

Ion 120.00 (119.70 to 120.70): V





#79

2-Chlorotoluene

Concen: 47.541 ug/l

RT: 11.73 min Scan# 1125

Delta R.T. -0.02 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :





VSTDCCC050

Tot Ion: 91 Resp: 425624

Ion Ratio Lower Upper

91 100

126 34.2 15.9 47.6



#80

1,3,5-Trimethylbenzene

Concen: 47.023 ug/l

RT: 11.83 min Scan# 1135

Delta R.T. -0.03 min

Lab File: VF060866.D




Acq: 29 Nov 2018 14:25


Tgt Ion: 105 Resp: 508967

Ion Ratio Lower Upper

105 100

120 48.1 25.8 77.3









#81
trans-1,4-Dichloro-2-butene
Concen: 65.569 ug/l
RT: 11.88 min Scan# 1140
Delta R.T. -0.02 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion: 75 Resp: 51240
Ion Ratio Lower Upper
75 100
53 74.8 65.9 98.9
89 51.6 38.8 58.2



#82
4-Chlorotoluene
Concen: 47.006 ug/l
RT: 11.91 min Scan# 1143
Delta R.T. -0.02 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25

Tgt Ion: 91 Resp: 442868
Ion Ratio Lower Upper
91 100
126 32.0 14.8 44.4









#83
tert-Butylbenzene
Concen: 47.959 ug/l
RT: 12.13 min Scan# 1166
Delta R.T. -0.03 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25




Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050


Tot Ion:119 Resp: 528772
Ion Ratio Lower Upper
119 100
91 60.9 30.1 90.3
134 24.6 12.6 37.8



#84
1,2,4-Trimethylbenzene
Concen: 48.303 ug/l
RT: 12.21 min Scan# 1174
Delta R.T. -0.02 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25

Tgt Ion:105 Resp: 521789
Ion Ratio Lower Upper
105 100
120 48.2 23.8 71.5









#85
 sec-Butylbenzene
 Concen: 46.696 ug/l
 RT: 12.31 min Scan# 1184
 Delta R.T. -0.02 min
 Lab File: VF060866.D
 Acq: 29 Nov 2018 14:25




Instrument :
 MSVOA_F
 ClientSampleId :
 VSTDCCC050


Tot Ion:105 Resp: 704785
 Ion Ratio Lower Upper
 105 100
 134 19.4 9.5 28.5



#86
 p-Isopropyltoluene
 Concen: 44.125 ug/l
 RT: 12.47 min Scan# 1200
 Delta R.T. -0.02 min
 Lab File: VF060866.D
 Acq: 29 Nov 2018 14:25

Tgt Ion:119 Resp: 558358
 Ion Ratio Lower Upper
 119 100
 134 26.1 12.9 38.6
 91 23.3 11.8 35.3





#87

1,3-Dichlorobenzene

Concen: 44.604 ug/l

RT: 12.48 min Scan# 1201

Delta R.T. -0.03 min





Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Tot Ion:146 Resp: 248004

Ion	Ratio	Lower	Upper
146	100		
111	41.0	20.0	59.9
148	64.1	32.9	98.6



#88

1,4-Dichlorobenzene

Concen: 45.495 ug/l

RT: 12.57 min Scan# 1210




Delta R.T. -0.03 min


Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Tgt Ion:146 Resp: 246516

Ion	Ratio	Lower	Upper
146	100		
111	42.2	18.9	56.7
148	66.8	32.3	96.9





#89

n-Butylbenzene

Concen: 47.978 ug/l

RT: 12.85 min Scan# 1239

Delta R.T. -0.02 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050





Tot Ion: 91 Resp: 613886

Ion Ratio Lower Upper

91 100

92 65.9 31.6 94.7

134 28.1 13.8 41.4



#90

Hexachloroethane

Concen: 46.993 ug/l

RT: 12.92 min Scan# 1246

Delta R.T. -0.02 min

Lab File: VF060866.D




Acq: 29 Nov 2018 14:25


Tgt Ion: 117 Resp: 143194

Ion Ratio Lower Upper

117 100

201 57.2 26.7 80.1





#91

1,2-Dichlorobenzene

Concen: 47.070 ug/l

RT: 12.95 min Scan# 1249

Delta R.T. -0.02 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion:146 Resp: 247841

Ion Ratio Lower Upper

146 100

111 39.9 21.2 63.6




148 64.9 32.5 97.5



Abundance Ion 146.00 (145.70 to 146.70): V

Ion 111.00 (110.70 to 111.70): V

Ion 148.00 (147.70 to 148.70): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 52.119 ug/l

RT: 13.65 min Scan# 1320

Delta R.T. -0.02 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25


Tgt Ion: 75 Resp: 22111

Ion Ratio Lower Upper

75 100

155 102.2 45.6 136.9



157 130.2 63.8 191.5




Abundance Ion 75.00 (74.70 to 75.70): VF0

Ion 155.00 (154.70 to 155.70): V

Ion 157.00 (156.70 to 157.70): V





#93

1,2,4-Trichlorobenzene

Concen: 48.489 ug/l

RT: 14.20 min Scan# 1376

Delta R.T. -0.02 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25

Instrument :

MSVOA_F

ClientSampleId :

VSTDCCC050


Tot Ion:180 Resp: 178394

Ion Ratio Lower Upper

180 100

182 93.7 48.5 145.5




145 37.4 18.1 54.3



Abundance Ion 180.00 (179.70 to 180.70): V

Ion 182.00 (181.70 to 182.70): V

Ion 145.00 (144.70 to 145.70): V



#94

Hexachlorobutadiene

Concen: 45.472 ug/l

RT: 14.19 min Scan# 1375

Delta R.T. -0.02 min

Lab File: VF060866.D

Acq: 29 Nov 2018 14:25


Tgt Ion:225 Resp: 111520

Ion Ratio Lower Upper

225 100

223 62.6 31.4 94.3



227 64.2 32.5 97.5




Abundance Ion 225.00 (224.70 to 225.70): V

Ion 223.00 (222.70 to 223.70): V

Ion 227.00 (226.70 to 227.70): V









#95
Naphthalene
Concen: 50.841 ug/l
RT: 14.46 min Scan# 1402
Delta R.T. -0.02 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25

Instrument :
MSVOA_F
ClientSampleId :
VSTDCCC050

Tot Ion:128 Resp: 362625
Ion Ratio Lower Upper
128 100
127 11.8 9.9 14.9
129 11.3 9.1 13.7



#96
1,2,3-Trichlorobenzene
Concen: 49.060 ug/l
RT: 14.61 min Scan# 1417
Delta R.T. -0.02 min
Lab File: VF060866.D
Acq: 29 Nov 2018 14:25

Tgt Ion:180 Resp: 167591
Ion Ratio Lower Upper
180 100
182 97.3 47.9 143.8
145 35.7 17.2 51.5

