

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF112918\
 Data File : VF060875.D
 Acq On : 29 Nov 2018 19:13
 Operator : VA/AP
 Sample : J5771-09
 Misc : 6.16g/5mL/MSVOA-F/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 EO-02-112918-B

Manual Integrations
 APPROVED

apatel
 12/3/2018 12:34:28 PM

Quant Time: Nov 30 06:28:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 27 01:36:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.87	168	192872	50.00	ug/l	-0.04
34) 1,4-Difluorobenzene	5.59	114	343922	50.00	ug/l	-0.04
63) Chlorobenzene-d5	9.78	117	309518	50.00	ug/l	-0.04
72) 1,4-Dichlorobenzene-d4	12.56	152	156411	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.84	65	129175	56.85	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	113.70%	
35) Dibromofluoromethane	4.11	113	145593	53.36	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	106.72%	
50) Toluene-d8	7.56	98	363149	45.18	ug/l	-0.04
Spiked Amount 50.000			Recovery	=	90.36%	
62) 4-Bromofluorobenzene	11.41	95	167002	45.72	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	91.44%	
Target Compounds						
20) Methylene Chloride	2.26	84	20886m	13.588	ug/l	Qvalue
52) Toluene	7.63	92	11357	1.919	ug/l	95
68) m/p-Xylenes	10.14	106	4741	1.176	ug/l	100
84) 1,2,4-Trimethylbenzene	12.21	105	11404	1.139	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF112918\
Data File : VF060875.D
Acq On : 29 Nov 2018 19:13
Operator : VA/AP
Sample : J5771-09
Misc : 6.16g/5mL/MSVOA-F/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
EO-02-112918-B

Manual Integrations
APPROVED

apatel
12/3/2018 12:34:28 PM

Quant Time: Nov 30 06:28:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 27 01:36:26 2018
Response via : Initial Calibration

