

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF120318\
 Data File : VF060893.D
 Acq On : 3 Dec 2018 16:02
 Operator : VA/AP
 Sample : J5771-11
 Misc : 5.81g/5mL/MSVOA-F/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 EO-02-112918-C

Manual Integrations
 APPROVED

apatel
 12/4/2018 2:49:22 PM

Quant Time: Dec 04 03:41:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 27 01:36:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.89	168	189990	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	5.62	114	350475	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.79	117	301147	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	12.57	152	142457	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.86	65	122088	54.55	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	109.10%	
35) Dibromofluoromethane	4.13	113	137636	49.50	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	99.00%	
50) Toluene-d8	7.57	98	307314	37.52	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	75.04%	
62) 4-Bromofluorobenzene	11.43	95	147988	39.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.52%	
Target Compounds						
20) Methylene Chloride	2.26	84	15321m	10.119	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF120318\
Data File : VF060893.D
Acq On : 3 Dec 2018 16:02
Operator : VA/AP
Sample : J5771-11
Misc : 5.81g/5mL/MSVOA-F/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
EO-02-112918-C

Manual Integrations
APPROVED

apatel
12/4/2018 2:49:22 PM

Quant Time: Dec 04 03:41:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 27 01:36:26 2018
Response via : Initial Calibration

