

Data Path : Z:\VOASRV\HPCHEM1\MSVOA F\DATA\VF120318\
 Data File : VF060894.D
 Acq On : 3 Dec 2018 16:30
 Operator : VA/AP
 Sample : J5764-23
 Misc : 6.64g/5mL/MSVOA-F/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_F
 ClientSampleId :
 JC-02-113018-B

Manual Integrations
 APPROVED

apatel
 12/4/2018 2:49:26 PM

Quant Time: Dec 04 03:54:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 27 01:36:26 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	4.88	168	188341	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	5.61	114	330321	50.00	ug/l	-0.02
63) Chlorobenzene-d5	9.80	117	292679	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	12.57	152	141044	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	4.87	65	112444	50.68	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	101.36%	
35) Dibromofluoromethane	4.13	113	119431	45.58	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	91.16%	
50) Toluene-d8	7.58	98	291505	37.76	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	75.52%	
62) 4-Bromofluorobenzene	11.42	95	138199	39.39	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	78.78%	
Target Compounds						
20) Methylene Chloride	2.27	84	16041m	10.687	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA F\DATA\VF120318\
Data File : VF060894.D
Acq On : 3 Dec 2018 16:30
Operator : VA/AP
Sample : J5764-23
Misc : 6.64g/5mL/MSVOA-F/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_F
ClientSampleId :
JC-02-113018-B

Manual Integrations
APPROVED

apatel
12/4/2018 2:49:26 PM

Quant Time: Dec 04 03:54:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_F\METHODS\82F112618S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 27 01:36:26 2018
Response via : Initial Calibration

